

BRAMMER STANDARD COMPANY, INC.

Provisional Certificate of Analysis BS 171C

Certified Reference Material for Cobalt Base Alloy Stellite 25 / L605 - UNS Number R30605

	Estimated Analysis ¹		Estimated Analysis ¹
Analysis listed as percent by weight			
Al	0.04	N	<0.05
B	<0.05	Nb	<0.05
C	0.12	Ni	10.4
Co	[50.5]	O	<0.005
Cr	20.4	P	0.003
Cu	0.03	S	0.0007
Fe	1.1	Si	0.05
La	0.05	Ta	<0.5
Mg	<0.005	Ti	0.073
Mn	1.5	V	0.013
Mo	0.13	W	15.3

¹ The estimated value listed is the present best estimate of the true value. Values are given in weight percent.

Form: This CRM is machined in the form of a disc approximately 38 mm in diameter and 19 mm thick by Brammer Standard Company, Inc.

A detailed final certificate of analysis will be supplied by January 8, 2021.

