



# National Institute of Standards & Technology

## Certificate of Analysis

### Standard Reference Material 697

#### Bauxite (Dominican)

(In Cooperation with the American Society for Testing and Materials)

This Standard Reference Material (SRM) is in the form of fine powder (nominal particle size < 0.08 mm) and is intended for use in evaluating chemical and instrumental methods of analyses.

(All analyses are based on samples dried 2 hours at 140 °C)

<u>Constituent</u>	<u>Certified Value<sup>1</sup></u> <u>Percent by Weight</u>	<u>Estimated</u> <u>Uncertainty<sup>2</sup></u>
Al <sub>2</sub> O <sub>3</sub>	45.8	0.2
Fe <sub>2</sub> O <sub>3</sub>	20.0	.2
SiO <sub>2</sub>	6.81	.07
TiO <sub>2</sub>	2.52	.05
ZrO <sub>2</sub>	0.065	.007
P <sub>2</sub> O <sub>5</sub>	.97	.06
V <sub>2</sub> O <sub>5</sub>	.063	.005
Cr <sub>2</sub> O <sub>3</sub>	.100	.005
CaO	.71	.03
MgO	.18	.02
MnO	.41	.03
ZnO	.037	.003
K <sub>2</sub> O	.062	.007
SO <sub>3</sub>	0.0770	0.0018
Loss on Ignition <sup>3</sup>	22.1	.2

<sup>1</sup>The certified value listed for a constituent is the present best estimate of the "true" value. The certified values are given as the oxide on an equivalent weight basis and assume stoichiometry in the form of the oxide listed.

<sup>2</sup>The estimated uncertainty listed for a constituent is based on judgment and represents an evaluation of the combined effects of method imprecision, possible systematic errors among methods, and material variability for samples 1.0 g or more. (No attempt was made to derive exact statistical measure of imprecision because several methods were involved in the determination of most constituents.)

<sup>3</sup>Determined by igniting to constant weight at 1050 °C.

Gaithersburg, MD 20899  
January 2, 1991  
(Revision of certificate dated 8-24-79)

William P. Reed, Acting Chief  
Standard Reference Materials Program

(over)

## ADDITIONAL INFORMATION ON THE COMPOSITION

Constituents other than those certified are present in this material as indicated below. They are not certified but are given as additional information on the composition.

Constituent	Concentration, Percent by weight	Constituent	Concentration, Percent by weight
BaO	(0.015)	Co	(0.0013)
Na <sub>2</sub> O	(0.036)	Hf	(0.0014)
Ce	(0.069)	Sc	(0.0058)

The mineralogical composition of SRM 697 is 15% kaolinite, 50% gibbsite, 10% boehmite, 20% hematite, and 5% anatase. It was determined by x-ray diffraction studies at the Geological Survey, U.S. Department of the Interior, Reston, VA, (J.W. Hosterman). These results are semiquantitative (to the nearest 5%).

## PLANNING, PREPARATION, TESTING, ANALYSIS:

The base material for this SRM was mined in the Dominican Republic and was provided by the Aluminum Company of America, Alcoa Technical Center, Pittsburgh, PA, through the courtesy of H.B. Hartman. It was processed (crushed, ground, sieved, and mixed) at the Colorado School of Mines Research Institute under a contract with the National Institute of Standards and Technology.

Homogeneity testing was performed at NIST by J.S. Maples and T.E. Gills using instrumental neutron activation analysis.

Cooperative analyses for certification were performed in the following laboratories:

- Aluminum Company of America, Alcoa Center, PA, R.C. Obbink.
- Aluminum Company of Canada, Ltd., Arvida Research Center, Arvida, Quebec, Canada, L. Girolami.
- Andrew S. McCreath & Son, Inc., Harrisburg, PA, F.A. Pennington, Jr., R.F. Eakin, and S.L. Miller.
- General Refractories Co., U.S. Refractories Division, Research Center, Baltimore, MD, S. Banerjee.
- Geological Survey, U.S. Department of the Interior, Reston, VA, H.J. Rose, Jr., and J.W. Hosterman.
- Kaiser Aluminum and Chemical Corp., Center for Technology, Pleasanton, CA, H.J. Seim, A.E. McLaughlin, D.F.G. Marten, A. Kermaninejad, R.C. Kinne, J.R. Skarset, J. Boruk, and U. Vogel.
- National Institute of Standards and Technology, Gaithersburg, MD, R.K. Bell (Retired), ASTM-NIST Assistant Research Associate W.R. Kelly and K.E. Murphy.
- National-Southwire Aluminum Co., Hawesville, KY, N. Robinson and E. Gotzy.
- Ormet Corp., Burnside, LA, W.L. Brown and A.D. Lafleur.
- Reynolds Aluminum Co., Alumina Research Division, Bauxite, AR, J.B. Ezell, Jr.
- University of Kentucky, Institute for Mining and Minerals Research, Center for Energy Research Laboratory, Lexington, KY, T.V. Rebagay.

The overall coordination of the technical measurements leading to certification were performed under the direction of J.I. Shultz, Research Associate, ASTM-NIST Research Associate Program.

The original, technical and support aspects involved in the preparation, certification, and issuance of this Standard Reference Material were coordinated through the Standard Reference Materials Program by R.E. Michaelis and R. Alvarez.

The update and revision of this Certificate of Analysis was coordinated through the Standard Reference Materials Program by T.E. Gills.