




Certificate

SILICIFIED MICROCRYSTALLINE CELLULOSE

(a mixture of microcrystalline cellulose and silicon dioxide)

USP Catalog No.: 1098399

USP Lot No.: R07550

<p>Microcrystalline Cellulose 98%</p>  <p>Silicon Dioxide 2%</p> 	CAS No.: n/a
	Molecular Formula: mixture
	Molecular Weight: n/a



LABEL TEXT

For use with specified USP compendial tests.
Not for use as a drug. See SDS prior to use
at www.usp.org/sds.

USP REFERENCE STANDARD
SILICIFIED MICROCRYSTALLINE CELLULOSE 1 g

Labeled for qualitative use. Keep container tightly
closed. Protect from moisture. This material is
hygroscopic.

USP, 12601 Twinbrook Pkwy, Rockville, MD, +1-301-881-0666
Cat. No. 1098399 Material mfd. in United States

LOT: R07550



Jeri L. Joth

Quality Assurance

**Assigned Value**

Please refer to the USP Reference Standard label for the assigned value of the specific lot. If the lot was developed for qualitative use, it will not have an assigned value.

Valid Use Date

It is the responsibility of the user to ascertain that a particular lot of a USP Reference Standard has official status either as a "Current lot" or as a "Previous lot" within the assigned valid use date. The online USP Reference Standards Catalog and the online USP Store at www.usp.org are updated daily. USP recommends referring to one of these sources prior to using a USP Reference Standard to make sure the lot is valid for use.

Storage

Storage conditions are lot-specific and may change from one lot to another. If no specific directions or limitations are provided on the USP Reference Standard label, the conditions of storage shall include storage at room temperature and protection from moisture, light, freezing, and excessive heat.

Instructions for Use

Follow the instructions provided on this Certificate, on the label of the USP Reference Standard, and in the associated USP documentary standard(s). Please refer to General Chapter <11> for additional information.

Non-USP Compendial Use

USP Reference Standards are for use in analytical or laboratory applications as specified in USP compendia. They are not for use in humans or animals as drugs or medical devices. It is the responsibility of the user to determine the suitability of the USP Reference Standard for non-USP compendial uses.

LEGAL NOTICE

USP WARRANTS GOOD TITLE TO USP REFERENCE STANDARDS ON DISPATCH FROM USP. THE FOREGOING WARRANTY IS IN LIEU OF ANY OTHER WARRANTIES, EXPRESS OR IMPLIED, INCLUDING WITHOUT LIMITATION ANY WARRANTY OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR PURPOSE, OR ANY WARRANTY THAT THE PRODUCTS, INCLUDING THIS CERTIFICATE, ARE OF MERCHANTABILITY QUALITY. USP'S LIABILITY ARISING OUT OF OR RELATING TO THE SUPPLY OF USP REFERENCE STANDARDS AND THIS CERTIFICATE SHALL IN NO EVENT INCLUDE LOSS OF PROFITS, COST OF PROCURING SUBSTITUTE GOODS OR SERVICES, OR ANY INCIDENTAL, INDIRECT, OR CONSEQUENTIAL DAMAGES OF ANY KIND, EVEN IF USP IS AWARE OF THE POSSIBILITY OF SUCH DAMAGES. WITHOUT LIMITING THE GENERALITY OF THE FOREGOING, USP DOES NOT WARRANT THAT THE USE OR RESALE OF USP REFERENCE STANDARDS, INCLUDING THEIR USE TO PERFORM TESTS AND ASSAYS PUBLISHED BY USP, WILL NOT INFRINGE UNITED STATES OR ANY OTHER PATENTS.

USP Reference Standards are not intended for use as drugs, dietary supplements, or as medical devices.

This certificate may not be reproduced without the express written permission of USP.