



EUROPEAN COMMISSION
JOINT RESEARCH CENTRE

Institute for Reference Materials and Measurements (Geel)

CERTIFIED REFERENCE MATERIAL BCR[®] – 286

CERTIFICATE OF ANALYSIS

ELECTROLYTICALLY REFINED LEAD			
	Mass fraction		Number of accepted sets of data p
	Certified value ¹⁾ [mg/kg]	Uncertainty ²⁾ [mg/kg]	
Bi	21.5	0.5	18
Sb	0.099	0.021	5
Tl	2.47	0.07	8
Zn	< 0.1 ³⁾		8
<p>1) Unweighted mean value of the means of p accepted sets of data, each set being obtained in a different laboratory and/or with a different method. The value is traceable to the International System of Units (SI).</p> <p>2) Half-width of the 95 % confidence interval of the mean defined in (1)</p> <p>3) The upper limit is supported by 8 sets of data, either actually measured values or upper limits. The true value lies below the stated value with a probability of 95 %.</p>			

This certificate is valid for five years after purchase.

Sales date:

The minimum amount of sample to be used is 1 g.

NOTE

This material has been certified by BCR (Community Bureau of Reference, the former reference materials programme of the European Commission). The certificate has been revised under the responsibility of IRMM.

Brussels, December 1984
Latest revision: August 2015

Signed: _____

Prof. Dr. Hendrik Emons
European Commission
Joint Research Centre
Institute for Reference Materials and Measurements
Retieseweg 111
B-2440 Geel, Belgium

Additional Material Information	
	Mass fraction
	Value ¹⁾ [µg/kg]
Ag	15
As	< 2
Cd	125
Cu	1490
Ni	41
Se	< 50
Sn	< 50
Te	< 100
1) Unweighted mean value of the means of accepted datasets. Each set was obtained in a different laboratory and/or with a different method. The value is traceable to the International System of Units (SI).	

DESCRIPTION OF THE SAMPLE

The material consists of electrolytically refined lead. The material is available in two forms:

BCR-286A: solid squares of about 60 mm x 60 mm x 12 mm

BCR-286B: chips in glass bottles containing about 160 g

ANALYTICAL METHOD USED FOR CERTIFICATION

Neutron activation analysis

Photon activation analysis

Flame atomic absorption spectrometry

Electrothermal atomic absorption spectrometry

Inductively coupled plasma emission spectrometry

Direct current plasma emission spectrometry

Laser enhanced ionisation (optogalvanic spectrometry)

Spectrophotometry

X-ray fluorescence

These techniques of final determination were preceded by dry or wet sample treatments as appropriate.

PARTICIPANTS

Accumulateurs Tudor, Archennes (BE)

Associated Lead Manufacturers, Research Laboratories, Greenford (GB)

Blei- und Silberhütte Braubach, Braubach (DE)

Bundesanstalt für Materialforschung und -prüfung, Berlin (DE)

Chloride Power Storage, Swinton (GB)

CNRS, Centre d'Etudes de Chimie Métallurgique, Vitry-sur-Seine (FR)

ENEA, Lab. Chimica Analitica, Roma (IT)

European Commission, Joint Research Centre, Ispra (IT)

Laboratoire National d'Essais, Paris (FR)

Metallurgie Hoboken-Overpelt, Hoboken (BE)

Minemet Recherche, Trappes (FR)

Norddeutsche Affinerie AG, Hamburg (DE)

Preussag AG Metall, Oker/Harz (DE)

Preussag Boliden Blei, Nordenham (DE)

Rijksuniversiteit Gent, Instituut voor Nucleaire Wetenschappen, Gent (BE)

Tonolli Grezzi (Sameton), Paderno Dugnano-Mi (IT)

Vieille Montagne, Angleur (BE)

SAFETY INFORMATION

The usual laboratory safety precautions apply.

INSTRUCTIONS FOR USE

The main purpose of the material is to assess method performance, i.e. checking accuracy of analytical results. As any reference material, it can also be used for control charts or validation studies.

Comparing an analytical result with the certified value

A result is unbiased if the combined uncertainty of measurement and certified value covers the difference between the certified value and the measurement result.

See also ERM Application Note 1; www.erm-crm.org

Use in quality control charts

The materials can be used for quality control charts. Different CRM units will give the same result as sufficient homogeneity has been demonstrated.

Use as a calibrant

It is not recommended to use matrix materials as calibrants. If used nevertheless, the uncertainty of the certified value shall be taken into consideration in the final estimation of measurement uncertainty.

When the solid material is used for spark emission spectrometry, the usual sample treatment (mechanical machining to have a fresh surface) should be applied. However, homogeneity for this low sample intake has not been demonstrated.

STORAGE

The material shall be stored at + 18 °C in the dark. However, the European Commission cannot be held responsible for changes that happen during storage of the material at the customer's premises, especially of opened samples.

LEGAL NOTICE

Neither IRMM, its subsidiaries, its contractors nor any person acting on their behalf,

(a) make any warranty or representation, express or implied that the use of any information, material, apparatus, method or process disclosed in this document does not infringe any privately owned intellectual property rights;

or

(b) assume any liability with respect to, or for damages resulting from, the use of any information, material, apparatus, method or process disclosed in this document save for loss or damage arising solely and directly from the negligence of IRMM or any of its subsidiaries.

NOTE

A technical report on the production of BCR-286 is available on the internet (<http://www.irmm.jrc.be>). A paper copy can be obtained from IRMM on request.

European Commission – Joint Research Centre
Institute for Reference Materials and Measurements (IRMM)
Retieseweg 111, 2440 Geel (Belgium)
Telephone: +32-(0)14-571.722 - Telefax: +32-(0)14-590.406