



Certificate of Analysis

Standard Reference Material[®] 1951c

Lipids in Frozen Human Serum

This Standard Reference Material (SRM) is intended primarily for use in evaluating the accuracy of clinical procedures for the determination of total cholesterol, high-density lipoprotein (HDL)-cholesterol, low-density lipoprotein (LDL)-cholesterol, and total glycerides in human serum. It is also intended for use in validating working or secondary reference materials. A unit of SRM 1951c consists of four vials of frozen human serum, two vials each at two different analyte concentration levels. Each bottle contains approximately 1 mL of serum.

Certified Concentration Values: The certified concentration values for total cholesterol and total glycerides (as triolein) are provided in Table 1a (mmol/L) and Table 1b (mg/dL). A NIST certified value is a value for which NIST has the highest confidence in its accuracy in that all known or suspected sources of bias have been investigated or taken into account [1]. The certified values for these analytes were determined at NIST and at the Lipid Reference Laboratory at the Centers for Disease Control and Prevention (CDC) by isotope dilution gas chromatography–mass spectrometry (ID-GC-MS) [2–5]. The methods used by NIST and CDC for measurement of total cholesterol and by NIST for measurement of total glycerides are recognized as higher-order reference measurement procedures by the Joint Committee for Traceability in Laboratory Medicine (JCTLM) [6]. The certified concentrations apply only to serum thawed to room temperature, 20 °C to 25 °C (see “Instructions for Storage and Use”).

Reference Values: The reference values for HDL-cholesterol, LDL-cholesterol, and total cholesterol, provided by the Lipid Reference Laboratory at CDC, are provided in Table 2. Reference values are noncertified values that are the best estimate of the true values based on available data; however, the values do not meet the NIST criteria for certification and are provided with associated uncertainties that may reflect only measurement precision, may not include all sources of uncertainty, or may reflect a lack of sufficient statistical agreement among multiple analytical methods [1]. These results were obtained using CDC reference methods for lipids [7,8].

Expiration of Certification: The certification of **SRM 1951c** is valid, within the measurement uncertainty specified, until **30 April 2018**, provided the SRM is handled and stored in accordance with instructions given in this certificate (see “Instructions for Storage and Use”). The certification is nullified if the SRM is damaged, contaminated, or otherwise modified.

Maintenance of Certification: NIST will monitor this SRM over the period of its certification. If substantive technical changes occur that affect the certification before expiration, NIST will notify the purchaser. Registration (see attached sheet) will facilitate notification.

The overall direction and coordination of the analytical measurements leading to the certification of this SRM were performed by K.W. Phinney of the NIST Biomolecular Measurement Division and L.C. Sander of the NIST Chemical Sciences Division. The overall coordination of the measurements at CDC was performed by H. Vesper.

Acquisition of the material was performed by K.W. Phinney and K.E. Sharpless of the NIST Chemical Sciences Division. The analytical measurements at NIST were performed by L.T. Sniegoski and M.J. Welch of the NIST Chemical Sciences Division. Analytical measurements at CDC were performed by S. Edwards, S. Pyatt, and S. Stribling of the CDC Division of Laboratory Sciences.

Carlos A. Gonzalez, Chief
Chemical Sciences Division

Gaithersburg, MD 20899
Certificate Issue Date: 27 June 2013

Robert L. Watters, Jr., Director
Office of Reference Materials

Statistical analysis was provided by N.-F. Zhang of the NIST Statistical Engineering Division.

Support aspects involved in the issuance of this SRM were coordinated through the NIST Office of Reference Materials.

WARNINGS TO USERS

Warning: SRM 1951c IS INTENDED FOR IN-VITRO DIAGNOSTIC USE ONLY. THIS IS A HUMAN-SOURCE MATERIAL. HANDLE PRODUCT AS A BIOHAZARDOUS MATERIAL CAPABLE OF TRANSMITTING INFECTIOUS DISEASE. The supplier of the serum has reported that each donor unit of serum used in the preparation of this product has been tested by an FDA-approved method and found non-reactive/negative for hepatitis B surface antigen (HbsAg), human immunodeficiency (HIV) 1 and 2 antibodies, and hepatitis C virus (HCV). However, no known test method can offer complete assurance that hepatitis B virus, hepatitis C virus, HIV, or other infectious agents are absent from this material. Accordingly, this human blood-based product should be handled at the Biosafety Level 2 or higher as recommended for any POTENTIALLY INFECTIOUS HUMAN SERUM OR BLOOD SPECIMEN in the Centers for Disease Control and Prevention/National Institutes of Health Manual [9].

INSTRUCTIONS FOR STORAGE AND USE

Storage: The serum is shipped frozen (on dry ice), and upon receipt, should be stored frozen until ready for use. A freezer temperature of $-20\text{ }^{\circ}\text{C}$ is acceptable for storage up to one week. If a longer storage time is anticipated, the material should be stored at or below $-60\text{ }^{\circ}\text{C}$. The SRM should not be exposed to sunlight or ultraviolet radiation. Storage of thawed material at room or refrigerator temperatures may result in changes in the analyte concentrations.

Use: Bottles of the SRM to be analyzed should be removed from the freezer and allowed to stand at room temperature until thawed. After the material is thawed to room temperature, it should be used **immediately**. The material should be swirled gently to mix it before aliquots are withdrawn.

SOURCE, PREPARATION, AND ANALYSIS⁽¹⁾

Source of Material: SRM 1951c was prepared by Solomon Park Research Laboratories (Kirkland, WA) following a protocol developed by the Cholesterol Reference Materials Subcommittee of the National Committee for Clinical Laboratory Standards (NCCLS) [10]. The goal of the NCCLS project was to develop a commutable lipid reference material for total cholesterol that would be useful in most presently available field methods. A large-scale study of a prior lot of this material involving most of the major clinical measurement systems found no significant biases between results on that prior lot and those from fresh, unpooled serum. The study verified that material prepared following the recommendations of the NCCLS study is an appropriate mechanism for transferring accuracy from the definitive and reference methods to the clinical laboratories without significant matrix effects on the systems tested.

Preparation of Material: Donor units were collected and allowed to clot at room temperature for 4 h. The serum was removed from the clot and immediately cooled to approximately $4\text{ }^{\circ}\text{C}$. Each unit of donor serum was then analyzed for total cholesterol content to determine which donor units to pool. The donor units selected were then pooled. One-milliliter aliquots of the bulk pool were dispensed into 3 mL glass vials and frozen at $-70\text{ }^{\circ}\text{C}$. This was accomplished within 50 hours of the initial donor unit collection.

Analysis: Value assignment of the concentrations of total cholesterol and total glycerides (as triolein) was based on the combination of results from ID-GC-MS methods at NIST and CDC. In addition, CDC provided results for total cholesterol using the Abell-Kendall reference method [7]. Concentrations of HDL-cholesterol were determined at CDC using the ultracentrifugation reference method [8], and LDL-cholesterol was determined at CDC using the beta-quantification reference method [8].

Homogeneity Analysis: The homogeneity assessment was made at the time the certification analyses were performed. A stratified sampling plan was devised to test for homogeneity across the lot of vials. There was no apparent trend in the data when plotted against the sequence in which the vials were prepared.

⁽¹⁾ Certain commercial equipment, instruments, or materials are identified in this certificate to adequately specify the experimental procedure. Such identification does not imply recommendation or endorsement by the National Institute of Standards and Technology, nor does it imply that the materials or equipment identified are necessarily the best available for the purpose.

Certified Values: Values for total cholesterol and total glycerides are based upon the means of the results from analyses at NIST and CDC using ID-GC-MS. The uncertainty provided with each certified value is an expanded uncertainty about the mean to cover the measurand with approximately 95 % confidence; it incorporates Type B uncertainty components related to the analyses and expresses both the observed difference between the results from the methods and their respective uncertainties, consistent with the ISO/JCGM Guide [11,12]. The expanded uncertainty is calculated as $U = ku_c$, where u_c is the combined uncertainty and k is a coverage factor corresponding to approximately 95 % confidence for each analyte [11]. For the certified values shown in Table 1, $k = 2$.

Table 1a. Certified Amount-of-Substance Concentration Values for SRM 1951c

Analyte	Level 1 (mmol/L)	Level 2 (mmol/L)
Total Cholesterol	3.943 ± 0.046	6.244 ± 0.072
Total Glycerides	1.717 ± 0.036	1.642 ± 0.036

Table 1b. Certified Mass Concentration Values for SRM 1951c

Analyte	Level 1 (mg/dL)	Level 2 (mg/dL)
Total Cholesterol	152.44 ± 1.78	241.41 ± 2.80
Total Glycerides ^(a)	152.0 ± 3.2	145.4 ± 3.2

^(a) Total glycerides results are expressed as triolein in milligrams per deciliter.

Reference Values: Values for HDL-cholesterol, LDL-cholesterol, and total cholesterol are based upon the means of results from analyses at CDC. The uncertainty provided with each reference value in Table 2 is an expanded uncertainty, U , at the 95 % level of confidence, and is calculated according to the method described in the ISO/JCGM Guide [11].

Table 2. Reference Values for HDL-Cholesterol, LDL-Cholesterol, and Total Cholesterol

Analyte	Level I (mg/dL)	Level II (mg/dL)
HDL-Cholesterol	41.0 ± 0.9	64.9 ± 1.7
LDL-Cholesterol	86.4 ± 1.4	143.8 ± 2.1
Total Cholesterol ^(a)	154.6 ± 1.1	244.8 ± 1.1

^(a) Total cholesterol as determined by the Abell-Kendall reference method [7].

REFERENCES

- [1] May, W.; Parris, R.; Beck, C.; Fassett, J.; Greenberg, R.; Guenther, F.; Kramer, G.; Wise, S.; Gills, T.; Colbert, J.; Gettings, R.; MacDonald, B.; *Definitions of Terms and Modes Used at NIST for Value-Assignment of Reference Materials for Chemical Measurements*; NIST Special Publication 260-136; U.S. Government Printing Office: Washington, DC (2000); available at <http://www.nist.gov/srm/publications.cfm> (accessed June 2013).
- [2] Ellerbe, P.; Meiselman, S.; Sniegowski, L.T.; Welch, M.J.; White, V.E.; *Determination of Serum Cholesterol by a Modification of the Isotope Dilution Mass Spectrometric Definitive Method*; *Anal. Chem.*, Vol. 61, pp. 1718–1723 (1989).
- [3] Edwards, S.H.; Kimberly, M.M.; Pyatt, S.D.; Stribling, S.L.; Dobbin, K.D.; Myers, G.L.; *Proposed Serum Cholesterol Reference Measurement Procedure by Gas Chromatography–Isotope Dilution Mass Spectrometry*; *Clin. Chem.*, Vol. 57, pp. 614–622 (2011).
- [4] Ellerbe, P.; Sniegowski, L.T.; Welch, M.J.; *Isotope Dilution Mass Spectrometry as a Candidate Definitive Method for Determining Total Glycerides and Triglycerides in Serum*; *Clin. Chem.*, Vol. 41, pp. 397–404 (1995).

- [5] Edwards, S.H.; Stribling, S.L.; Pyatt, S.D.; Kimberly, M.M.; *Reference Measurement Procedure for Total Glycerides by Isotope Dilution GC-MS*; Clin. Chem., Vol. 58, pp. 768–776 (2012).
- [6] Joint Committee for Traceability in Laboratory Medicine; available at <http://www.bipm.org/en/committees/jc/jctlm/> (accessed June 2013).
- [7] Abell, L.L.; Levy, B.B.; Brodie, B.B.; Kendall, F.E.; *A Simplified Method for the Determination of Total Cholesterol in Serum and Demonstration of Its Specificity*; J. Biol. Chem., Vol. 195, pp. 357–366 (1952).
- [8] Myers, G.L.; Cooper, G.R.; Greenberg, N.; Kimberly, M.M.; Waymack, P.P.; Hassemer, D.J.; *Standardization of Lipid and Lipoprotein Measurements*; In *Handbook of Lipoprotein Testing*, 2nd ed., Rifai, N.; Warnick, G.R.; Dominiczak, M.H., Eds.; American Association for Clinical Chemistry (AACC), Washington, D.C., pp. 717–748.
- [9] CDC/NIH; *Biosafety in Microbiological and Biomedical Laboratories, 5th ed.*; Richardson, J.; Barkley, W.E.; Richmond, J.; McKinney, R.W., Eds.; U.S. Department of Health and Human Services, Public Health Service, Centers for Disease Control and Prevention and National Institutes of Health; US Government Printing Office: Washington, D.C. (2009); available at http://www.cdc.gov/OD/OHS/biosfty/bmb15/BMBL_5th_Edition.pdf (accessed June 2013).
- [10] NCCLS; *Preparation and Validation of Commutable Frozen Human Serum Pools as Secondary Reference Materials for Cholesterol Measurement Procedures*; NCCLS Publication C37-A, National Committee for Clinical Laboratory Standards, Wayne, PA (2000).
- [11] JCGM 100:2008; *Evaluation of Measurement Data — Guide to the Expression of Uncertainty in Measurement* (GUM 1995 with Minor Corrections); Joint Committee for Guides in Metrology (2008); available at http://www.bipm.org/utis/common/documents/jcgm/JCGM_100_2008_E.pdf (accessed June 2013); see also Taylor, B.N.; Kuyatt, C.E.; *Guidelines for Evaluating and Expressing the Uncertainty of NIST Measurement Results*; NIST Technical Note 1297; U.S. Government Printing Office: Washington, DC (1994); available at <http://www.nist.gov/pml/pubs/tn1297/index.cfm> (accessed June 2013).
- [12] JCGM 101:2008; *Evaluation of Measurement Data – Supplement 1 to the Guide to the Expression of Uncertainty in Measurement – Propagation of Distributions Using a Monte Carlo Method*; Joint Committee for Guides in Metrology (JCGM) (2008); available at http://www.bipm.org/utis/common/documents/jcgm/JCGM_101_2008_E.pdf (accessed June 2013).

Users of this SRM should ensure that the Certificate of Analysis in their possession is current. This can be accomplished by contacting the SRM Program: telephone (301) 975-2200; fax (301) 948-3730; e-mail srminfo@nist.gov; or via the Internet at <http://www.nist.gov/srm>.