

# THE INSTITUTE FOR CERTIFIED REFERENCE MATERIALS ICRM

## CERTIFICATE OF ANALYSIS

### CERTIFIED REFERENCE MATERIAL

No. F28/2

## FERROPHOSPHORUS

CERTIFIED VALUES AND UNCERTAINTIES (95% confidence level), %

Phosphorus	16.05	$\pm 0.03$
Silicon	1.11	$\pm 0.02$
Manganese	1.20	$\pm 0.01$
Sulphur	0.021	$\pm 0.001$

ADDITIONAL DATA: Minimum weight of sample for analysis is 0.1 g.

Analytical methods used are given in Supplement.

Valeriy Stepanovskikh  
Director, ICRM

