

Tel./Fax: ++42155633 6834



**Producer:** *pb-anal, Hlinkova 25, 04001 Košice, Slovakia*  
*Institute of Radioecology, Komenského 9, 04001 Košice, Slovakia*

## **CERTIFICATE**

### **REFERENCE MATERIAL**

#### **SLUDGE FROM CITY WATER TREATMENT – WT-H**

Issued in accordance with § 7 of the Slovak Act No. 142/2000, Coll.

Registration No.: 105/07

**Sort of material** *Sludge from city water treatment – WT-H*  
**Code:** *12-3-14*

Certified values of individual major elements contents with expanded uncertainties (for k = 2)

Element	Mass fraction [%]	U [%]
Al	2.51	0.25
Ca	4.83	0.28
Fe	1.70	0.17
K	0.582	0.056

Element	Mass fraction [%]	U [%]
Mg	0.570	0.054
P	1.44	0.08
S	1.34	0.13

Certified values of individual minor elements contents with expanded uncertainties (for k = 2)

Element	Mass fraction [µg/g]	U [µg/g]
Ag	38.4	4.2
As	146	17
Ba	772	55
Be	0.68	0.11
Cd	54.7	4.2
Cl	6 170	450
Co	315	34
Cr	1340	160
Cu	3 140	310
Hg	31.3	2.5

Element	Mass fraction [µg/g]	U [µg/g]
Li	11.5	1.7
Mn	3 660	290
Mo	78.4	6.0
Ni	1 140	110
Pb	2 290	110
Sb	43.0	7.5
Sn	20.3	2.8
Sr	872	77
V	33.7	3.3
Zn	6 360	330

**Relationship:** Interlaboratory comparison by ISO Guide 35: 1989

**Packing:** (40±5) g

**Minimal sample size:** 200 mg

**Charge:** 12-3-14-10507

**Validity:** November 2016

**Authority responsible for preparation and certification :** RNDr. Štefan Bartha

In Bratislava, 20.9.2007



*Ing. Viliam Pátoprstý, PhD.*  
*Director of Chemical Centre*

**Certification conditions:** For certification of the reference material a method of inter-laboratory comparison by ISO Guide 35 was used. The certified values are expressed as average values from repeated measurements at least in 8 laboratories using at least 2 different analytical techniques, with the elimination of outlying results. Uncertainties are expressed in form of expanded uncertainty with uncertainty coefficient  $k=2$ .

**Properties:** Powder material, size of particles <0.08 mm

**Homogeneity:** CRM homogeneity is guaranteed for weights >100 mg

**Storage:** Keep in a dark, dry and cool room. After opening it, keep a well closed bottle in refrigerator under the temperature of 3-5°C.

**Expected use:** Verification of the correctness of analytical procedures employed at the analysis of samples with similar matrixes, validation of measurement methods.

**Instruction for use:** Prior to each use, the humidity in respective part of the reference material is determined by drying at temperature of 105°C to reach the constant mass. The rest of the reference material is treated according to requirements of the analytical procedure in use (ignition, converting into solution, extraction, briquetting/caking ...).

#### Informative values of mass fraction of different elements

Informative values of the mass fraction of individual elements and their standard uncertainties:

Element	Mass fraction [ $\mu\text{g/g}$ ]	$U$ [ $\mu\text{g/g}$ ]
B	31.0	8.0
Bi	3.0	0.6
Se	13.2	2.6

\* mass fraction in ng/g

Element	Mass fraction [%]	$U$ [%]
Na	0.282	0.075
Si	8.2	1.2

$U$  - expanded uncertainty

The informative values are expressed as average values from repeated measurements at least in 5 laboratories using at least 2 different analytical techniques, with the elimination of outlying results. Uncertainties are expressed in form of expanded uncertainty with uncertainty coefficient  $k=2$ .




Informative values without uncertainties:

Element	Mass fraction [ $\mu\text{g/g}$ ]
Au	0.9
Br	10
Ce	30
Cs	2
Eu	0.4
Hf	2
La	15
Lu	0.08
Nd	10
Rb	35

Element	Mass fraction [ $\mu\text{g/g}$ ]
Sc	3
Sm	2
Ta	0.5
Tb	0.3
Th	4
Ti	1 100
U	2
W	10
Yb	0.8
Zr	50

#### Participating laboratories:

1. Institut jadernych issledovanij i jadernoj energetiky, Sofia, Bulgaria
2. Ecole Polytechnique Montreal, Montreal, Canada
3. ČSAV, Ústav geológie a geotechniky, Praha, Czechoslovakia
4. Ústav rádioekológie a využitia jadrovej techniky, Košice, Czechoslovakia
5. Farmaceutická fakulta, Katedra analytickej chémie, Bratislava, Czechoslovakia
6. Výzkumný ústav energetický, Praha, Czechoslovakia
7. Výzkumný ústav stavebních hmot, Brno, Czechoslovakia
8. Geologický průzkum n. p., Brno, Czechoslovakia
9. Ústřední laboratoř ČSUP, Stráž pod Ralskem, Czechoslovakia
10. Prírodovedencká fakulta UK, Katedra analytickej chémie, Bratislava, Czechoslovakia
11. Ústav jadrového výzkumu, Řež u Prahy, Czechoslovakia
12. Palivový kombinát, 25. února k.p., Vřesová, Czechoslovakia
13. Fyzikální ústav ČSAV, Praha, Czechoslovakia
14. Výzkumný ústav pro hnědé uhlí, Most, Czechoslovakia
15. Oravské ferozliatinárské závody, n.p., Istebné, Czechoslovakia
16. VŠB, Katedra geológie a mineralógie, Ostrava-Poruba, Czechoslovakia
17. Geologický prieskum, n.p., Spišská Nová Ves, Czechoslovakia
18. Akademie der Wissenschaften der DDR, Zentralinstitut für Isotopen und Strahlenforschung, Leipzig, DDR
19. Humboldt-Universität zu Berlin „Bereich Medizin (Charite), Berlin, DDR
20. Institut für Physik der Universität Hohenheim, Stuttgart, Federal Republic of Germany
21. Bayerisches Landesamt für Umweltschutz, München, Federal Republic of Germany
22. Niedersächsisches Landesamt für Immissionsschutz, Institut für Arbeitsmedizin, Hannover, Federal Republic of Germany
23. Fachbereich Anorganische Chemie und Kerchemie, Technische Hochschule Darmstadt, Darmstadt, Federal Republic of Germany
24. Helsinki University of Technology, Department of Chemistry, Espoo, Finland
25. Environmental and Food Laboratory of Helsinki, Helsinki, Finland
26. National Public Health Institute, Helsinki, Finland
27. Technical University of Budapest, Training Reactor, Budapest, Hungary
28. Hungarian Geological Survey, Budapest, Hungary
29. Government of India, Air Monitoring Section, Bhabha Atomic Research Centre, Bombay, India
30. Institute for Physics and Nuclear Engineering, Bucharest, Romania
31. US Department of Energy, Environmental Measurements Lab., New York, U.S.A.



- 32. Institut gorjučich iskopaemykh, Moskva, USSR
- 33. Physics Institute, Latvian SSR Academy of Sciences, Riga, USSR
- 34. Istitut jadernoj fiziki AN UzSSR, Taškent, USSR
  
- 35. Slovenská geológia š.p., Spišská Nová Ves, Slovakia
- 36. Ekologické laboratória, Spišská Nová Ves, Slovakia
- 37. Oravské ferozliatinárske závody, Istebné, Slovakia
- 38. Geoekologické laboratória š.p., Turčianske Teplice, Slovakia
- 39. pb-anal, Košice, Slovakia
  
- 40. Geoanalytické laboratóriá, Spišská Nová Ves, Slovakia
- 41. pb-anal, Košice, Slovakia
- 42. VVDVÚ, Košice, Slovakia

*Appendix: Laboratories No. 1 - 34 are participants from intercomparison analyses in 1985 year.*

*Laboratories No. 35 - 39 are participants from intercomparison analyses in 1996 year.*

*Laboratories No. 40 - 42 are participants from intercomparison analyses in 2006 year.*

**Used analytical methods:**

**INSTRUMENTAL NEUTRON ACTIVATION ANALYSIS - GENERAL**

- INAA – short irradiation
- INAA – long irradiation
- INAA – with radioisotope separation

**ATOMIC ABSORPTION SPECTROMETRY - GENERAL**

- AAS – flame technique
- AAS – flameless technique
- AAS – hydride technique
- AAS – could vapour techniques
- AAS – other methods (TMA, AMA)

**ATOMIC EMISSION SPECTROMETRY - GENERAL**

- AES – flame technique
- AES – ICP
- AES – ICP + MS
- AES – other methods

**X – RAY SPECTROMETRY - GENERAL**

- RFA - energy dispersive
- RFA – wavelength dispersive

**MOLECULAR SPECTROMETRY - GENERAL**

- Spectrometry UV-VIS

**MASS SPECTROMETRY - GENERAL**

- MS – spark source

**ELECTROCHEMICAL METHODS - GENERAL**

- DPP
- Another

**CHEMICAL METHODS - GENERAL**

- Gravimetric analysis
- Volumetric analysis

**OTHER METHODS**



**Recertification:** CRM re-certification was performed in year 2006, on the base of stability confirmation of previously certified values. This confirmation is based on inter-laboratory analyses (by 3 different labs and 4 different measuring techniques) of Pb, Cu, Mn, K and Mg.

**Analytical technique:**

AAS – flame technique  
RFA – energy dispersive  
DPP

**Participants:**

Geoanalytické laboratóriá, Spišská Nová Ves, SLOVAKIA  
pb-anal, Lomená 1, Košice, SLOVAKIA  
VVDVÚ, Kukučínova 2, Košice, SLOVAKIA

Commission for Certified Reference Materials (CCRM) of the Slovak Republic on the base of obtained results and in accordance with its No. 1/2007 from its meeting decided to prolong the expire date of those CRM to 30/11/2016.

**In case of deterioration of the reference material by the user, the manufacturer is relieved from the liability to guarantee the certified values!**



**Distributed by:**

Labmix24 GmbH  
Jonas-Elkan-Weg 4  
46499 Hamminkeln, GERMANY  
Phone: +49 2856 90924-0  
Fax: +49 2856 90924-24  
[info@labmix24.com](mailto:info@labmix24.com)  
[www.labmix24.com](http://www.labmix24.com)

