

**Out of Stock - Item is not available at this time - Relative Intensity Correction Standard for Raman Spectroscopy Using 785 nm Excitation**

Art. ID                      NIST-2241  
Unit                        each  
Delivery details           No Dangerous Good /not restricted

**Description**

This Standard Reference Material (SRM®) is a certified spectroscopic standard for the correction of the relative intensity of Raman spectra obtained with instruments employing 785 nm laser excitation. NIST-2241 consists of an optical glass that emits a broadband luminescence spectrum when excited at this laser wavelength. The relative spectral intensity of the glass luminescence has been determined through the use of a white-light, uniform-source, integrating sphere that has been calibrated for its irradiance at NIST. The shape of the mean luminescence spectrum of this glass is described by a mathematical expression that relates the relative spectral intensity to the wavenumber (cm<sup>-1</sup>) expressed as the Raman shift from the excitation laser wavelength. This model, together with a measurement of the luminescence spectrum of the standard, can be used to determine the spectral intensity-response correction that is unique to each Raman system. The resulting instrument-intensity-response correction may then be used to obtain Raman spectra that are largely free from instrument-induced spectral artifacts. /// Sample value(s) - please ask for current certificate.

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
A0	Polynomial Coefficient (20 °C - 25 °C)		9,71937 E-02			
A1	Polynomial Coefficient (20 °C - 25 °C)		2,28325 E-04			
A2	Polynomial Coefficient (20 °C - 25 °C)		(-5,86762 E-08)			
A3	Polynomial Coefficient (20 °C - 25 °C)		2,16023 E-10			
A4	Polynomial Coefficient (20 °C - 25 °C)		(-9,77171 E-14)			
A5	Polynomial Coefficient (20 °C - 25 °C)		1,15596 E-17			