

## Lubricating Oil Calibration Curve

Art. ID                    ACCU-ASTM-P-0111-SET  
Unit                        17 x 100 mL  
Deliverydetails        No Dangerous Good /not restricted

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
ACCU-ASTM-P-0111-01	Barium (Ba)	[7440-39-3]	0,1	%w/w		
ACCU-ASTM-P-0111-01	Calcium (Ca)	[7440-70-2]	0,6	%w/w		
ACCU-ASTM-P-0111-01	Chlorine (Cl)	[7782-50-5]	0,1	%w/w		
ACCU-ASTM-P-0111-01	Phosphorus (P)	[7723-14-0]	0,005	%w/w		
ACCU-ASTM-P-0111-01	Sulfur (S)	[7704-34-9]	0,175	%w/w		
ACCU-ASTM-P-0111-01	Zinc (Zn)	[7440-66-6]	0,06	%w/w		
ACCU-ASTM-P-0111-02	Barium (Ba)	[7440-39-3]	0,175	%w/w		
ACCU-ASTM-P-0111-02	Calcium (Ca)	[7440-70-2]	0,5	%w/w		
ACCU-ASTM-P-0111-02	Chlorine (Cl)	[7782-50-5]	0	%w/w		
ACCU-ASTM-P-0111-02	Phosphorus (P)	[7723-14-0]	0,2	%w/w		
ACCU-ASTM-P-0111-02	Sulfur (S)	[7704-34-9]	0,05	%w/w		
ACCU-ASTM-P-0111-02	Zinc (Zn)	[7440-66-6]	0,08	%w/w		
ACCU-ASTM-P-0111-03	Barium (Ba)	[7440-39-3]	0	%w/w		
ACCU-ASTM-P-0111-03	Calcium (Ca)	[7440-70-2]	0,4	%w/w		
ACCU-ASTM-P-0111-03	Chlorine (Cl)	[7782-50-5]	0,01	%w/w		
ACCU-ASTM-P-0111-03	Phosphorus (P)	[7723-14-0]	0,15	%w/w		
ACCU-ASTM-P-0111-03	Sulfur (S)	[7704-34-9]	0,3	%w/w		
ACCU-ASTM-P-0111-03	Zinc (Zn)	[7440-66-6]	0,18	%w/w		
ACCU-ASTM-P-0111-04	Barium (Ba)	[7440-39-3]	0,025	%w/w		
ACCU-ASTM-P-0111-04	Calcium (Ca)	[7440-70-2]	0,26	%w/w		
ACCU-ASTM-P-0111-04	Chlorine (Cl)	[7782-50-5]	0,5	%w/w		
ACCU-ASTM-P-0111-04	Phosphorus (P)	[7723-14-0]	0,25	%w/w		
ACCU-ASTM-P-0111-04	Sulfur (S)	[7704-34-9]	0,15	%w/w		
ACCU-ASTM-P-0111-04	Zinc (Zn)	[7440-66-6]	0,12	%w/w		
ACCU-ASTM-P-0111-05	Barium (Ba)	[7440-39-3]	0,15	%w/w		
ACCU-ASTM-P-0111-05	Calcium (Ca)	[7440-70-2]	0,005	%w/w		
ACCU-ASTM-P-0111-05	Chlorine (Cl)	[7782-50-5]	1	%w/w		
ACCU-ASTM-P-0111-05	Phosphorus (P)	[7723-14-0]	0,005	%w/w		
ACCU-ASTM-P-0111-05	Sulfur (S)	[7704-34-9]	0,45	%w/w		
ACCU-ASTM-P-0111-05	Zinc (Zn)	[7440-66-6]	0,07	%w/w		
ACCU-ASTM-P-0111-06	Barium (Ba)	[7440-39-3]	0	%w/w		

ACCU-ASTM-P-0111-06	Calcium (Ca)	[7440-70-2]	0,4	%w/w
ACCU-ASTM-P-0111-06	Chlorine (Cl)	[7782-50-5]	0,4	%w/w
ACCU-ASTM-P-0111-06	Phosphorus (P)	[7723-14-0]	0,025	%w/w
ACCU-ASTM-P-0111-06	Sulfur (S)	[7704-34-9]	0,35	%w/w
ACCU-ASTM-P-0111-06	Zinc (Zn)	[7440-66-6]	0,1	%w/w
ACCU-ASTM-P-0111-07	Barium (Ba)	[7440-39-3]	0,2	%w/w
ACCU-ASTM-P-0111-07	Calcium (Ca)	[7440-70-2]	0,3	%w/w
ACCU-ASTM-P-0111-07	Chlorine (Cl)	[7782-50-5]	0,1	%w/w
ACCU-ASTM-P-0111-07	Phosphorus (P)	[7723-14-0]	0,06	%w/w
ACCU-ASTM-P-0111-07	Sulfur (S)	[7704-34-9]	0,25	%w/w
ACCU-ASTM-P-0111-07	Zinc (Zn)	[7440-66-6]	0,12	%w/w
ACCU-ASTM-P-0111-08	Barium (Ba)	[7440-39-3]	0	%w/w
ACCU-ASTM-P-0111-08	Calcium (Ca)	[7440-70-2]	0,2	%w/w
ACCU-ASTM-P-0111-08	Chlorine (Cl)	[7782-50-5]	0,01	%w/w
ACCU-ASTM-P-0111-08	Phosphorus (P)	[7723-14-0]	0,1	%w/w
ACCU-ASTM-P-0111-08	Sulfur (S)	[7704-34-9]	0,45	%w/w
ACCU-ASTM-P-0111-08	Zinc (Zn)	[7440-66-6]	0,1	%w/w
ACCU-ASTM-P-0111-09	Barium (Ba)	[7440-39-3]	0,1	%w/w
ACCU-ASTM-P-0111-09	Calcium (Ca)	[7440-70-2]	0,06	%w/w
ACCU-ASTM-P-0111-09	Chlorine (Cl)	[7782-50-5]	0,05	%w/w
ACCU-ASTM-P-0111-09	Phosphorus (P)	[7723-14-0]	0,08	%w/w
ACCU-ASTM-P-0111-09	Sulfur (S)	[7704-34-9]	0,3	%w/w
ACCU-ASTM-P-0111-09	Zinc (Zn)	[7440-66-6]	0,13	%w/w
ACCU-ASTM-P-0111-10	Barium (Ba)	[7440-39-3]	0,05	%w/w
ACCU-ASTM-P-0111-10	Calcium (Ca)	[7440-70-2]	0,06	%w/w
ACCU-ASTM-P-0111-10	Chlorine (Cl)	[7782-50-5]	0,2	%w/w
ACCU-ASTM-P-0111-10	Phosphorus (P)	[7723-14-0]	0,05	%w/w
ACCU-ASTM-P-0111-10	Sulfur (S)	[7704-34-9]	0,2	%w/w
ACCU-ASTM-P-0111-10	Zinc (Zn)	[7440-66-6]	0,05	%w/w
ACCU-ASTM-P-0111-11	Barium (Ba)	[7440-39-3]	0,075	%w/w
ACCU-ASTM-P-0111-11	Calcium (Ca)	[7440-70-2]	0,05	%w/w
ACCU-ASTM-P-0111-11	Chlorine (Cl)	[7782-50-5]	0,5	%w/w
ACCU-ASTM-P-0111-11	Phosphorus (P)	[7723-14-0]	0,12	%w/w
ACCU-ASTM-P-0111-11	Sulfur (S)	[7704-34-9]	0,1	%w/w
ACCU-ASTM-P-0111-11	Zinc (Zn)	[7440-66-6]	0,075	%w/w
ACCU-ASTM-P-0111-12	Barium (Ba)	[7440-39-3]	0,01	%w/w
ACCU-ASTM-P-0111-12	Calcium (Ca)	[7440-70-2]	0,025	%w/w
ACCU-ASTM-P-0111-12	Chlorine (Cl)	[7782-50-5]	0,8	%w/w

ACCU-ASTM-P-0111-12	Phosphorus (P)	[7723-14-0]	0,15	%w/w
ACCU-ASTM-P-0111-12	Sulfur (S)	[7704-34-9]	0,2	%w/w
ACCU-ASTM-P-0111-12	Zinc (Zn)	[7440-66-6]	0,13	%w/w
ACCU-ASTM-P-0111-13	Barium (Ba)	[7440-39-3]	0,005	%w/w
ACCU-ASTM-P-0111-13	Calcium (Ca)	[7440-70-2]	0,005	%w/w
ACCU-ASTM-P-0111-13	Chlorine (Cl)	[7782-50-5]	1	%w/w
ACCU-ASTM-P-0111-13	Phosphorus (P)	[7723-14-0]	0,2	%w/w
ACCU-ASTM-P-0111-13	Sulfur (S)	[7704-34-9]	0,4	%w/w
ACCU-ASTM-P-0111-13	Zinc (Zn)	[7440-66-6]	0,15	%w/w
ACCU-ASTM-P-0111-14	Barium (Ba)	[7440-39-3]	0	%w/w
ACCU-ASTM-P-0111-14	Calcium (Ca)	[7440-70-2]	0,17	%w/w
ACCU-ASTM-P-0111-14	Chlorine (Cl)	[7782-50-5]	0,6	%w/w
ACCU-ASTM-P-0111-14	Phosphorus (P)	[7723-14-0]	0,25	%w/w
ACCU-ASTM-P-0111-14	Sulfur (S)	[7704-34-9]	0,55	%w/w
ACCU-ASTM-P-0111-14	Zinc (Zn)	[7440-66-6]	0,11	%w/w
ACCU-ASTM-P-0111-15	Barium (Ba)	[7440-39-3]	0	%w/w
ACCU-ASTM-P-0111-15	Calcium (Ca)	[7440-70-2]	0,1	%w/w
ACCU-ASTM-P-0111-15	Chlorine (Cl)	[7782-50-5]	0,2	%w/w
ACCU-ASTM-P-0111-15	Phosphorus (P)	[7723-14-0]	0,1	%w/w
ACCU-ASTM-P-0111-15	Sulfur (S)	[7704-34-9]	0,2	%w/w
ACCU-ASTM-P-0111-15	Zinc (Zn)	[7440-66-6]	0,2	%w/w
ACCU-ASTM-P-0111-17	Barium (Ba)	[7440-39-3]	0	%w/w
ACCU-ASTM-P-0111-17	Calcium (Ca)	[7440-70-2]	0	%w/w
ACCU-ASTM-P-0111-17	Chlorine (Cl)	[7782-50-5]	0	%w/w
ACCU-ASTM-P-0111-17	Phosphorus (P)	[7723-14-0]	0	%w/w
ACCU-ASTM-P-0111-17	Sulfur (S)	[7704-34-9]	0	%w/w
ACCU-ASTM-P-0111-17	Zinc (Zn)	[7440-66-6]	0	%w/w