

Lubricating Oil Calibration Curve

Art. ID ACCU-ASTM-P-0113-SET
Unit 17 x 100 mL
Deliverydetails No Dangerous Good /not restricted

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
ACCU-ASTM-P-0113-01	Barium (Ba)	[7440-39-3]	0,025	%w/w		
ACCU-ASTM-P-0113-01	Calcium (Ca)	[7440-70-2]	0,6	%w/w		
ACCU-ASTM-P-0113-01	Magnesium (Mg)	[7439-95-4]	0,1	%w/w		
ACCU-ASTM-P-0113-01	Phosphorus (P)	[7723-14-0]	0,005	%w/w		
ACCU-ASTM-P-0113-01	Sulfur (S)	[7704-34-9]	0,175	%w/w		
ACCU-ASTM-P-0113-01	Zinc (Zn)	[7440-66-6]	0,06	%w/w		
ACCU-ASTM-P-0113-02	Barium (Ba)	[7440-39-3]	0	%w/w		
ACCU-ASTM-P-0113-02	Calcium (Ca)	[7440-70-2]	0,5	%w/w		
ACCU-ASTM-P-0113-02	Magnesium (Mg)	[7439-95-4]	0,15	%w/w		
ACCU-ASTM-P-0113-02	Phosphorus (P)	[7723-14-0]	0,2	%w/w		
ACCU-ASTM-P-0113-02	Sulfur (S)	[7704-34-9]	0,05	%w/w		
ACCU-ASTM-P-0113-02	Zinc (Zn)	[7440-66-6]	0,08	%w/w		
ACCU-ASTM-P-0113-03	Barium (Ba)	[7440-39-3]	0,1	%w/w		
ACCU-ASTM-P-0113-03	Calcium (Ca)	[7440-70-2]	0,4	%w/w		
ACCU-ASTM-P-0113-03	Magnesium (Mg)	[7439-95-4]	0,35	%w/w		
ACCU-ASTM-P-0113-03	Phosphorus (P)	[7723-14-0]	0,15	%w/w		
ACCU-ASTM-P-0113-03	Sulfur (S)	[7704-34-9]	0,3	%w/w		
ACCU-ASTM-P-0113-03	Zinc (Zn)	[7440-66-6]	0,18	%w/w		
ACCU-ASTM-P-0113-04	Barium (Ba)	[7440-39-3]	0,175	%w/w		
ACCU-ASTM-P-0113-04	Calcium (Ca)	[7440-70-2]	0,26	%w/w		
ACCU-ASTM-P-0113-04	Magnesium (Mg)	[7439-95-4]	0,225	%w/w		
ACCU-ASTM-P-0113-04	Phosphorus (P)	[7723-14-0]	0,25	%w/w		
ACCU-ASTM-P-0113-04	Sulfur (S)	[7704-34-9]	0,15	%w/w		
ACCU-ASTM-P-0113-04	Zinc (Zn)	[7440-66-6]	0,12	%w/w		
ACCU-ASTM-P-0113-05	Barium (Ba)	[7440-39-3]	0,15	%w/w		
ACCU-ASTM-P-0113-05	Calcium (Ca)	[7440-70-2]	0,005	%w/w		
ACCU-ASTM-P-0113-05	Magnesium (Mg)	[7439-95-4]	0,45	%w/w		
ACCU-ASTM-P-0113-05	Phosphorus (P)	[7723-14-0]	0,005	%w/w		
ACCU-ASTM-P-0113-05	Sulfur (S)	[7704-34-9]	0,45	%w/w		
ACCU-ASTM-P-0113-05	Zinc (Zn)	[7440-66-6]	0,07	%w/w		
ACCU-ASTM-P-0113-06	Barium (Ba)	[7440-39-3]	0	%w/w		

ACCU-ASTM-P-0113-06	Calcium (Ca)	[7440-70-2]	0,4	%w/w
ACCU-ASTM-P-0113-06	Magnesium (Mg)	[7439-95-4]	0,5	%w/w
ACCU-ASTM-P-0113-06	Phosphorus (P)	[7723-14-0]	0,025	%w/w
ACCU-ASTM-P-0113-06	Sulfur (S)	[7704-34-9]	0,35	%w/w
ACCU-ASTM-P-0113-06	Zinc (Zn)	[7440-66-6]	0,1	%w/w
ACCU-ASTM-P-0113-07	Barium (Ba)	[7440-39-3]	0,1	%w/w
ACCU-ASTM-P-0113-07	Calcium (Ca)	[7440-70-2]	0,3	%w/w
ACCU-ASTM-P-0113-07	Magnesium (Mg)	[7439-95-4]	0,325	%w/w
ACCU-ASTM-P-0113-07	Phosphorus (P)	[7723-14-0]	0,06	%w/w
ACCU-ASTM-P-0113-07	Sulfur (S)	[7704-34-9]	0,25	%w/w
ACCU-ASTM-P-0113-07	Zinc (Zn)	[7440-66-6]	0,12	%w/w
ACCU-ASTM-P-0113-08	Barium (Ba)	[7440-39-3]	0,2	%w/w
ACCU-ASTM-P-0113-08	Calcium (Ca)	[7440-70-2]	0,2	%w/w
ACCU-ASTM-P-0113-08	Magnesium (Mg)	[7439-95-4]	0,25	%w/w
ACCU-ASTM-P-0113-08	Phosphorus (P)	[7723-14-0]	0,1	%w/w
ACCU-ASTM-P-0113-08	Sulfur (S)	[7704-34-9]	0,45	%w/w
ACCU-ASTM-P-0113-08	Zinc (Zn)	[7440-66-6]	0,1	%w/w
ACCU-ASTM-P-0113-09	Barium (Ba)	[7440-39-3]	0,05	%w/w
ACCU-ASTM-P-0113-09	Calcium (Ca)	[7440-70-2]	0,06	%w/w
ACCU-ASTM-P-0113-09	Magnesium (Mg)	[7439-95-4]	0,1	%w/w
ACCU-ASTM-P-0113-09	Phosphorus (P)	[7723-14-0]	0,08	%w/w
ACCU-ASTM-P-0113-09	Sulfur (S)	[7704-34-9]	0,3	%w/w
ACCU-ASTM-P-0113-09	Zinc (Zn)	[7440-66-6]	0,13	%w/w
ACCU-ASTM-P-0113-10	Barium (Ba)	[7440-39-3]	0,075	%w/w
ACCU-ASTM-P-0113-10	Calcium (Ca)	[7440-70-2]	0,06	%w/w
ACCU-ASTM-P-0113-10	Magnesium (Mg)	[7439-95-4]	0,4	%w/w
ACCU-ASTM-P-0113-10	Phosphorus (P)	[7723-14-0]	0,05	%w/w
ACCU-ASTM-P-0113-10	Sulfur (S)	[7704-34-9]	0,2	%w/w
ACCU-ASTM-P-0113-10	Zinc (Zn)	[7440-66-6]	0,05	%w/w
ACCU-ASTM-P-0113-11	Barium (Ba)	[7440-39-3]	0,01	%w/w
ACCU-ASTM-P-0113-11	Calcium (Ca)	[7440-70-2]	0,05	%w/w
ACCU-ASTM-P-0113-11	Magnesium (Mg)	[7439-95-4]	0,3	%w/w
ACCU-ASTM-P-0113-11	Phosphorus (P)	[7723-14-0]	0,12	%w/w
ACCU-ASTM-P-0113-11	Sulfur (S)	[7704-34-9]	0,1	%w/w
ACCU-ASTM-P-0113-11	Zinc (Zn)	[7440-66-6]	0,075	%w/w
ACCU-ASTM-P-0113-12	Barium (Ba)	[7440-39-3]	0	%w/w
ACCU-ASTM-P-0113-12	Calcium (Ca)	[7440-70-2]	0,025	%w/w
ACCU-ASTM-P-0113-12	Magnesium (Mg)	[7439-95-4]	0,2	%w/w

ACCU-ASTM-P-0113-12	Phosphorus (P)	[7723-14-0]	0,15	%w/w
ACCU-ASTM-P-0113-12	Sulfur (S)	[7704-34-9]	0,2	%w/w
ACCU-ASTM-P-0113-12	Zinc (Zn)	[7440-66-6]	0,13	%w/w
ACCU-ASTM-P-0113-13	Barium (Ba)	[7440-39-3]	0,175	%w/w
ACCU-ASTM-P-0113-13	Calcium (Ca)	[7440-70-2]	0,005	%w/w
ACCU-ASTM-P-0113-13	Magnesium (Mg)	[7439-95-4]	0,375	%w/w
ACCU-ASTM-P-0113-13	Phosphorus (P)	[7723-14-0]	0,2	%w/w
ACCU-ASTM-P-0113-13	Sulfur (S)	[7704-34-9]	0,4	%w/w
ACCU-ASTM-P-0113-13	Zinc (Zn)	[7440-66-6]	0,15	%w/w
ACCU-ASTM-P-0113-14	Barium (Ba)	[7440-39-3]	0,005	%w/w
ACCU-ASTM-P-0113-14	Calcium (Ca)	[7440-70-2]	0,17	%w/w
ACCU-ASTM-P-0113-14	Magnesium (Mg)	[7439-95-4]	0,175	%w/w
ACCU-ASTM-P-0113-14	Phosphorus (P)	[7723-14-0]	0,25	%w/w
ACCU-ASTM-P-0113-14	Sulfur (S)	[7704-34-9]	0,55	%w/w
ACCU-ASTM-P-0113-14	Zinc (Zn)	[7440-66-6]	0,11	%w/w
ACCU-ASTM-P-0113-15	Barium (Ba)	[7440-39-3]	0	%w/w
ACCU-ASTM-P-0113-15	Calcium (Ca)	[7440-70-2]	0,1	%w/w
ACCU-ASTM-P-0113-15	Magnesium (Mg)	[7439-95-4]	0,425	%w/w
ACCU-ASTM-P-0113-15	Phosphorus (P)	[7723-14-0]	0,1	%w/w
ACCU-ASTM-P-0113-15	Sulfur (S)	[7704-34-9]	0,2	%w/w
ACCU-ASTM-P-0113-15	Zinc (Zn)	[7440-66-6]	0,2	%w/w
ACCU-ASTM-P-0113-16	Barium (Ba)	[7440-39-3]	0,005	%w/w
ACCU-ASTM-P-0113-16	Calcium (Ca)	[7440-70-2]	0,01	%w/w
ACCU-ASTM-P-0113-16	Magnesium (Mg)	[7439-95-4]	0,275	%w/w
ACCU-ASTM-P-0113-16	Phosphorus (P)	[7723-14-0]	0,01	%w/w
ACCU-ASTM-P-0113-16	Sulfur (S)	[7704-34-9]	0,6	%w/w
ACCU-ASTM-P-0113-16	Zinc (Zn)	[7440-66-6]	0,25	%w/w
ACCU-ASTM-P-0113-17	Barium (Ba)	[7440-39-3]		
ACCU-ASTM-P-0113-17	Calcium (Ca)	[7440-70-2]		
ACCU-ASTM-P-0113-17	Magnesium (Mg)	[7439-95-4]		
ACCU-ASTM-P-0113-17	Phosphorus (P)	[7723-14-0]		
ACCU-ASTM-P-0113-17	Sulfur (S)	[7704-34-9]		
ACCU-ASTM-P-0113-17	Zinc (Zn)	[7440-66-6]		