

Lubricating Oil Calibration Curve

Art. ID ACCU-ASTM-P-0117-SET
Unit 10 x 100 mL
Deliverydetails No Dangerous Good /not restricted

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
ACCU-ASTM-P-0117-01	Calcium (Ca)	[7440-70-2]	0,02	%w/w		
ACCU-ASTM-P-0117-01	Chlorine (Cl)	[7782-50-5]	0,03	%w/w		
ACCU-ASTM-P-0117-01	Copper (Cu)	[7440-50-8]	0,01	%w/w		
ACCU-ASTM-P-0117-01	Magnesium (Mg)	[7439-95-4]	0,2	%w/w		
ACCU-ASTM-P-0117-01	Phosphorus (P)	[7723-14-0]	0,25	%w/w		
ACCU-ASTM-P-0117-01	Sulfur (S)	[7704-34-9]	1	%w/w		
ACCU-ASTM-P-0117-01	Zinc (Zn)	[7440-66-6]	0,02	%w/w		
ACCU-ASTM-P-0117-02	Calcium (Ca)	[7440-70-2]	0,02	%w/w		
ACCU-ASTM-P-0117-02	Chlorine (Cl)	[7782-50-5]	0,02	%w/w		
ACCU-ASTM-P-0117-02	Copper (Cu)	[7440-50-8]	0,05	%w/w		
ACCU-ASTM-P-0117-02	Magnesium (Mg)	[7439-95-4]	0,2	%w/w		
ACCU-ASTM-P-0117-02	Phosphorus (P)	[7723-14-0]	0,02	%w/w		
ACCU-ASTM-P-0117-02	Sulfur (S)	[7704-34-9]	0,02	%w/w		
ACCU-ASTM-P-0117-02	Zinc (Zn)	[7440-66-6]	0,25	%w/w		
ACCU-ASTM-P-0117-03	Calcium (Ca)	[7440-70-2]	0,02	%w/w		
ACCU-ASTM-P-0117-03	Chlorine (Cl)	[7782-50-5]	0,2	%w/w		
ACCU-ASTM-P-0117-03	Copper (Cu)	[7440-50-8]	0,01	%w/w		
ACCU-ASTM-P-0117-03	Magnesium (Mg)	[7439-95-4]	0,04	%w/w		
ACCU-ASTM-P-0117-03	Phosphorus (P)	[7723-14-0]	0,25	%w/w		
ACCU-ASTM-P-0117-03	Sulfur (S)	[7704-34-9]	0,15	%w/w		
ACCU-ASTM-P-0117-03	Zinc (Zn)	[7440-66-6]	0,25	%w/w		
ACCU-ASTM-P-0117-04	Calcium (Ca)	[7440-70-2]	0,02	%w/w		
ACCU-ASTM-P-0117-04	Chlorine (Cl)	[7782-50-5]	0,2	%w/w		
ACCU-ASTM-P-0117-04	Copper (Cu)	[7440-50-8]	0,05	%w/w		
ACCU-ASTM-P-0117-04	Magnesium (Mg)	[7439-95-4]	0,04	%w/w		
ACCU-ASTM-P-0117-04	Phosphorus (P)	[7723-14-0]	0,02	%w/w		
ACCU-ASTM-P-0117-04	Sulfur (S)	[7704-34-9]	1	%w/w		
ACCU-ASTM-P-0117-04	Zinc (Zn)	[7440-66-6]	0,02	%w/w		
ACCU-ASTM-P-0117-05	Calcium (Ca)	[7440-70-2]	0,4	%w/w		
ACCU-ASTM-P-0117-05	Chlorine (Cl)	[7782-50-5]	0,02	%w/w		
ACCU-ASTM-P-0117-05	Copper (Cu)	[7440-50-8]	0,01	%w/w		

ACCU-ASTM-P-0117-05	Magnesium (Mg)	[7439-95-4]	0,04	%w/w
ACCU-ASTM-P-0117-05	Phosphorus (P)	[7723-14-0]	0,02	%w/w
ACCU-ASTM-P-0117-05	Sulfur (S)	[7704-34-9]	1	%w/w
ACCU-ASTM-P-0117-05	Zinc (Zn)	[7440-66-6]	0,25	%w/w
ACCU-ASTM-P-0117-06	Calcium (Ca)	[7440-70-2]	0,4	%w/w
ACCU-ASTM-P-0117-06	Chlorine (Cl)	[7782-50-5]	0,02	%w/w
ACCU-ASTM-P-0117-06	Copper (Cu)	[7440-50-8]	0,05	%w/w
ACCU-ASTM-P-0117-06	Magnesium (Mg)	[7439-95-4]	0,04	%w/w
ACCU-ASTM-P-0117-06	Phosphorus (P)	[7723-14-0]	0,25	%w/w
ACCU-ASTM-P-0117-06	Sulfur (S)	[7704-34-9]	0,02	%w/w
ACCU-ASTM-P-0117-06	Zinc (Zn)	[7440-66-6]	0,02	%w/w
ACCU-ASTM-P-0117-07	Calcium (Ca)	[7440-70-2]	0,4	%w/w
ACCU-ASTM-P-0117-07	Chlorine (Cl)	[7782-50-5]	0,2	%w/w
ACCU-ASTM-P-0117-07	Copper (Cu)	[7440-50-8]	0,01	%w/w
ACCU-ASTM-P-0117-07	Magnesium (Mg)	[7439-95-4]	0,2	%w/w
ACCU-ASTM-P-0117-07	Phosphorus (P)	[7723-14-0]	0,02	%w/w
ACCU-ASTM-P-0117-07	Sulfur (S)	[7704-34-9]	0,02	%w/w
ACCU-ASTM-P-0117-07	Zinc (Zn)	[7440-66-6]	0,05	%w/w
ACCU-ASTM-P-0117-08	Calcium (Ca)	[7440-70-2]	0,4	%w/w
ACCU-ASTM-P-0117-08	Chlorine (Cl)	[7782-50-5]	0,2	%w/w
ACCU-ASTM-P-0117-08	Copper (Cu)	[7440-50-8]	0,05	%w/w
ACCU-ASTM-P-0117-08	Magnesium (Mg)	[7439-95-4]	0,2	%w/w
ACCU-ASTM-P-0117-08	Phosphorus (P)	[7723-14-0]	0,25	%w/w
ACCU-ASTM-P-0117-08	Sulfur (S)	[7704-34-9]	1	%w/w
ACCU-ASTM-P-0117-08	Zinc (Zn)	[7440-66-6]	0,25	%w/w
ACCU-ASTM-P-0117-09	Calcium (Ca)	[7440-70-2]	0,2	%w/w
ACCU-ASTM-P-0117-09	Chlorine (Cl)	[7782-50-5]	0,1	%w/w
ACCU-ASTM-P-0117-09	Copper (Cu)	[7440-50-8]	0,025	%w/w
ACCU-ASTM-P-0117-09	Magnesium (Mg)	[7439-95-4]	0,08	%w/w
ACCU-ASTM-P-0117-09	Phosphorus (P)	[7723-14-0]	0,15	%w/w
ACCU-ASTM-P-0117-09	Sulfur (S)	[7704-34-9]	0,5	%w/w
ACCU-ASTM-P-0117-09	Zinc (Zn)	[7440-66-6]	0,1	%w/w
ACCU-ASTM-P-0117-10	Calcium (Ca)	[7440-70-2]	0	%w/w
ACCU-ASTM-P-0117-10	Copper (Cu)	[7440-50-8]	0	%w/w
ACCU-ASTM-P-0117-10	Magnesium (Mg)	[7439-95-4]	0	%w/w
ACCU-ASTM-P-0117-10	Phosphorus (P)	[7723-14-0]	0	%w/w
ACCU-ASTM-P-0117-10	Sulfur (S)	[7704-34-9]	0	%w/w
ACCU-ASTM-P-0117-10	Sulfur (S)	[7704-34-9]	0	%w/w

ACCU-ASTM-P-0117-10	Zinc (Zn)	[7440-66-6]	0	%w/w
---------------------	-----------	-------------	---	------