

Lubricating Oil Calibration Curve

Art. ID	ACCU-ASTM-P-0118-SET
Unit	10 x 100 mL
Deliverydetails	No Dangerous Good /not restricted

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
ACCU-ASTM-P-0118-01	Barium (Ba)	[7440-39-3]	0,02	%w/w		
ACCU-ASTM-P-0118-01	Calcium (Ca)	[7440-70-2]	0,02	%w/w		
ACCU-ASTM-P-0118-01	Chlorine (Cl)	[7782-50-5]	0,03	%w/w		
ACCU-ASTM-P-0118-01	Copper (Cu)	[7440-50-8]	0,01	%w/w		
ACCU-ASTM-P-0118-01	Magnesium (Mg)	[7439-95-4]	0,2	%w/w		
ACCU-ASTM-P-0118-01	Phosphorus (P)	[7723-14-0]	0,25	%w/w		
ACCU-ASTM-P-0118-01	Sulfur (S)	[7704-34-9]	1	%w/w		
ACCU-ASTM-P-0118-01	Zinc (Zn)	[7440-66-6]	0,02	%w/w		
ACCU-ASTM-P-0118-02	Barium (Ba)	[7440-39-3]	0,25	%w/w		
ACCU-ASTM-P-0118-02	Calcium (Ca)	[7440-70-2]	0,02	%w/w		
ACCU-ASTM-P-0118-02	Chlorine (Cl)	[7782-50-5]	0,02	%w/w		
ACCU-ASTM-P-0118-02	Copper (Cu)	[7440-50-8]	0,05	%w/w		
ACCU-ASTM-P-0118-02	Magnesium (Mg)	[7439-95-4]	0,2	%w/w		
ACCU-ASTM-P-0118-02	Phosphorus (P)	[7723-14-0]	0,02	%w/w		
ACCU-ASTM-P-0118-02	Sulfur (S)	[7704-34-9]	0,02	%w/w		
ACCU-ASTM-P-0118-02	Zinc (Zn)	[7440-66-6]	0,25	%w/w		
ACCU-ASTM-P-0118-03	Barium (Ba)	[7440-39-3]	0,02	%w/w		
ACCU-ASTM-P-0118-03	Calcium (Ca)	[7440-70-2]	0,02	%w/w		
ACCU-ASTM-P-0118-03	Chlorine (Cl)	[7782-50-5]	0,2	%w/w		
ACCU-ASTM-P-0118-03	Copper (Cu)	[7440-50-8]	0,01	%w/w		
ACCU-ASTM-P-0118-03	Magnesium (Mg)	[7439-95-4]	0,04	%w/w		
ACCU-ASTM-P-0118-03	Phosphorus (P)	[7723-14-0]	0,25	%w/w		
ACCU-ASTM-P-0118-03	Sulfur (S)	[7704-34-9]	0,15	%w/w		
ACCU-ASTM-P-0118-03	Zinc (Zn)	[7440-66-6]	0,25	%w/w		
ACCU-ASTM-P-0118-04	Barium (Ba)	[7440-39-3]	0,25	%w/w		
ACCU-ASTM-P-0118-04	Calcium (Ca)	[7440-70-2]	0,02	%w/w		
ACCU-ASTM-P-0118-04	Chlorine (Cl)	[7782-50-5]	0,2	%w/w		
ACCU-ASTM-P-0118-04	Copper (Cu)	[7440-50-8]	0,05	%w/w		
ACCU-ASTM-P-0118-04	Magnesium (Mg)	[7439-95-4]	0,04	%w/w		
ACCU-ASTM-P-0118-04	Phosphorus (P)	[7723-14-0]	0,02	%w/w		
ACCU-ASTM-P-0118-04	Sulfur (S)	[7704-34-9]	1	%w/w		

ACCU-ASTM-P-0118-04	Zinc (Zn)	[7440-66-6]	0,02	%w/w
ACCU-ASTM-P-0118-05	Barium (Ba)	[7440-39-3]	0,02	%w/w
ACCU-ASTM-P-0118-05	Calcium (Ca)	[7440-70-2]	0,4	%w/w
ACCU-ASTM-P-0118-05	Chlorine (Cl)	[7782-50-5]	0,02	%w/w
ACCU-ASTM-P-0118-05	Copper (Cu)	[7440-50-8]	0,01	%w/w
ACCU-ASTM-P-0118-05	Magnesium (Mg)	[7439-95-4]	0,04	%w/w
ACCU-ASTM-P-0118-05	Phosphorus (P)	[7723-14-0]	0,02	%w/w
ACCU-ASTM-P-0118-05	Sulfur (S)	[7704-34-9]	1	%w/w
ACCU-ASTM-P-0118-05	Zinc (Zn)	[7440-66-6]	0,25	%w/w
ACCU-ASTM-P-0118-06	Barium (Ba)	[7440-39-3]	0,25	%w/w
ACCU-ASTM-P-0118-06	Calcium (Ca)	[7440-70-2]	0,4	%w/w
ACCU-ASTM-P-0118-06	Chlorine (Cl)	[7782-50-5]	0,02	%w/w
ACCU-ASTM-P-0118-06	Copper (Cu)	[7440-50-8]	0,05	%w/w
ACCU-ASTM-P-0118-06	Magnesium (Mg)	[7439-95-4]	0,04	%w/w
ACCU-ASTM-P-0118-06	Phosphorus (P)	[7723-14-0]	0,25	%w/w
ACCU-ASTM-P-0118-06	Sulfur (S)	[7704-34-9]	0,02	%w/w
ACCU-ASTM-P-0118-06	Zinc (Zn)	[7440-66-6]	0,02	%w/w
ACCU-ASTM-P-0118-07	Barium (Ba)	[7440-39-3]	0,02	%w/w
ACCU-ASTM-P-0118-07	Calcium (Ca)	[7440-70-2]	0,4	%w/w
ACCU-ASTM-P-0118-07	Chlorine (Cl)	[7782-50-5]	0,2	%w/w
ACCU-ASTM-P-0118-07	Copper (Cu)	[7440-50-8]	0,01	%w/w
ACCU-ASTM-P-0118-07	Magnesium (Mg)	[7439-95-4]	0,2	%w/w
ACCU-ASTM-P-0118-07	Phosphorus (P)	[7723-14-0]	0,02	%w/w
ACCU-ASTM-P-0118-07	Sulfur (S)	[7704-34-9]	0,02	%w/w
ACCU-ASTM-P-0118-07	Zinc (Zn)	[7440-66-6]	0,05	%w/w
ACCU-ASTM-P-0118-08	Barium (Ba)	[7440-39-3]	0,25	%w/w
ACCU-ASTM-P-0118-08	Calcium (Ca)	[7440-70-2]	0,4	%w/w
ACCU-ASTM-P-0118-08	Chlorine (Cl)	[7782-50-5]	0,2	%w/w
ACCU-ASTM-P-0118-08	Copper (Cu)	[7440-50-8]	0,05	%w/w
ACCU-ASTM-P-0118-08	Magnesium (Mg)	[7439-95-4]	0,2	%w/w
ACCU-ASTM-P-0118-08	Phosphorus (P)	[7723-14-0]	0,25	%w/w
ACCU-ASTM-P-0118-08	Sulfur (S)	[7704-34-9]	1	%w/w
ACCU-ASTM-P-0118-08	Zinc (Zn)	[7440-66-6]	0,25	%w/w
ACCU-ASTM-P-0118-09	Barium (Ba)	[7440-39-3]	0,13	%w/w
ACCU-ASTM-P-0118-09	Calcium (Ca)	[7440-70-2]	0,2	%w/w
ACCU-ASTM-P-0118-09	Chlorine (Cl)	[7782-50-5]	0,1	%w/w
ACCU-ASTM-P-0118-09	Copper (Cu)	[7440-50-8]	0,025	%w/w
ACCU-ASTM-P-0118-09	Magnesium (Mg)	[7439-95-4]	0,08	%w/w

ACCU-ASTM-P-0118-09	Phosphorus (P)	[7723-14-0]	0,15	%w/w
ACCU-ASTM-P-0118-09	Sulfur (S)	[7704-34-9]	0,5	%w/w
ACCU-ASTM-P-0118-09	Zinc (Zn)	[7440-66-6]	0,1	%w/w
ACCU-ASTM-P-0118-10	Barium (Ba)	[7440-39-3]	0	%w/w
ACCU-ASTM-P-0118-10	Calcium (Ca)	[7440-70-2]	0	%w/w
ACCU-ASTM-P-0118-10	Chlorine (Cl)	[7782-50-5]	0	%w/w
ACCU-ASTM-P-0118-10	Copper (Cu)	[7440-50-8]	0	%w/w
ACCU-ASTM-P-0118-10	Magnesium (Mg)	[7439-95-4]	0	%w/w
ACCU-ASTM-P-0118-10	Phosphorus (P)	[7723-14-0]	0	%w/w
ACCU-ASTM-P-0118-10	Sulfur (S)	[7704-34-9]	0	%w/w
ACCU-ASTM-P-0118-10	Zinc (Zn)	[7440-66-6]	0	%w/w