

Lubricating Oil Calibration Curve

Art. ID ACCU-ASTM-P-0119-SET
Unit 22 x 100 mL
Deliverydetails No Dangerous Good /not restricted

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
ACCU-ASTM-P-0119-01	Calcium (Ca)	[7440-70-2]	0,3	%w/w		
ACCU-ASTM-P-0119-01	Chlorine (Cl)	[7782-50-5]	0,08	%w/w		
ACCU-ASTM-P-0119-01	Copper (Cu)	[7440-50-8]	0,03	%w/w		
ACCU-ASTM-P-0119-01	Magnesium (Mg)	[7439-95-4]	0,06	%w/w		
ACCU-ASTM-P-0119-01	Phosphorus (P)	[7723-14-0]	0,06	%w/w		
ACCU-ASTM-P-0119-01	Sulfur (S)	[7704-34-9]	0,275	%w/w		
ACCU-ASTM-P-0119-01	Zinc (Zn)	[7440-66-6]	0,06	%w/w		
ACCU-ASTM-P-0119-02	Calcium (Ca)	[7440-70-2]	0,25	%w/w		
ACCU-ASTM-P-0119-02	Chlorine (Cl)	[7782-50-5]	0,1	%w/w		
ACCU-ASTM-P-0119-02	Copper (Cu)	[7440-50-8]	0	%w/w		
ACCU-ASTM-P-0119-02	Magnesium (Mg)	[7439-95-4]	0,01	%w/w		
ACCU-ASTM-P-0119-02	Phosphorus (P)	[7723-14-0]	0,15	%w/w		
ACCU-ASTM-P-0119-02	Sulfur (S)	[7704-34-9]	0	%w/w		
ACCU-ASTM-P-0119-02	Zinc (Zn)	[7440-66-6]	0,15	%w/w		
ACCU-ASTM-P-0119-03	Calcium (Ca)	[7440-70-2]	0,5	%w/w		
ACCU-ASTM-P-0119-03	Chlorine (Cl)	[7782-50-5]	0	%w/w		
ACCU-ASTM-P-0119-03	Copper (Cu)	[7440-50-8]	0,035	%w/w		
ACCU-ASTM-P-0119-03	Magnesium (Mg)	[7439-95-4]	0,16	%w/w		
ACCU-ASTM-P-0119-03	Phosphorus (P)	[7723-14-0]	0,15	%w/w		
ACCU-ASTM-P-0119-03	Sulfur (S)	[7704-34-9]	0	%w/w		
ACCU-ASTM-P-0119-03	Zinc (Zn)	[7440-66-6]	0,02	%w/w		
ACCU-ASTM-P-0119-04	Calcium (Ca)	[7440-70-2]	0,35	%w/w		
ACCU-ASTM-P-0119-04	Chlorine (Cl)	[7782-50-5]	0,01	%w/w		
ACCU-ASTM-P-0119-04	Copper (Cu)	[7440-50-8]	0	%w/w		
ACCU-ASTM-P-0119-04	Magnesium (Mg)	[7439-95-4]	0,12	%w/w		
ACCU-ASTM-P-0119-04	Phosphorus (P)	[7723-14-0]	0,08	%w/w		
ACCU-ASTM-P-0119-04	Sulfur (S)	[7704-34-9]	0,2	%w/w		
ACCU-ASTM-P-0119-04	Zinc (Zn)	[7440-66-6]	0	%w/w		
ACCU-ASTM-P-0119-05	Calcium (Ca)	[7440-70-2]	0,11	%w/w		
ACCU-ASTM-P-0119-05	Chlorine (Cl)	[7782-50-5]	0	%w/w		
ACCU-ASTM-P-0119-05	Copper (Cu)	[7440-50-8]	0,015	%w/w		

ACCU-ASTM-P-0119-05	Magnesium (Mg)	[7439-95-4]	0,1	%w/w
ACCU-ASTM-P-0119-05	Phosphorus (P)	[7723-14-0]	0,1	%w/w
ACCU-ASTM-P-0119-05	Sulfur (S)	[7704-34-9]	0,3	%w/w
ACCU-ASTM-P-0119-05	Zinc (Zn)	[7440-66-6]	0,05	%w/w
ACCU-ASTM-P-0119-06	Calcium (Ca)	[7440-70-2]	0,2	%w/w
ACCU-ASTM-P-0119-06	Chlorine (Cl)	[7782-50-5]	0,1	%w/w
ACCU-ASTM-P-0119-06	Copper (Cu)	[7440-50-8]	0	%w/w
ACCU-ASTM-P-0119-06	Magnesium (Mg)	[7439-95-4]	0,2	%w/w
ACCU-ASTM-P-0119-06	Phosphorus (P)	[7723-14-0]	0,05	%w/w
ACCU-ASTM-P-0119-06	Sulfur (S)	[7704-34-9]	0,25	%w/w
ACCU-ASTM-P-0119-06	Zinc (Zn)	[7440-66-6]	0,15	%w/w
ACCU-ASTM-P-0119-07	Calcium (Ca)	[7440-70-2]	0	%w/w
ACCU-ASTM-P-0119-07	Chlorine (Cl)	[7782-50-5]	0,05	%w/w
ACCU-ASTM-P-0119-07	Copper (Cu)	[7440-50-8]	0,025	%w/w
ACCU-ASTM-P-0119-07	Magnesium (Mg)	[7439-95-4]	0	%w/w
ACCU-ASTM-P-0119-07	Phosphorus (P)	[7723-14-0]	0	%w/w
ACCU-ASTM-P-0119-07	Sulfur (S)	[7704-34-9]	0,45	%w/w
ACCU-ASTM-P-0119-07	Zinc (Zn)	[7440-66-6]	0,02	%w/w
ACCU-ASTM-P-0119-08	Calcium (Ca)	[7440-70-2]	0,15	%w/w
ACCU-ASTM-P-0119-08	Chlorine (Cl)	[7782-50-5]	0,03	%w/w
ACCU-ASTM-P-0119-08	Copper (Cu)	[7440-50-8]	0	%w/w
ACCU-ASTM-P-0119-08	Magnesium (Mg)	[7439-95-4]	0,1	%w/w
ACCU-ASTM-P-0119-08	Phosphorus (P)	[7723-14-0]	0,03	%w/w
ACCU-ASTM-P-0119-08	Sulfur (S)	[7704-34-9]	0,4	%w/w
ACCU-ASTM-P-0119-08	Zinc (Zn)	[7440-66-6]	0,04	%w/w
ACCU-ASTM-P-0119-09	Calcium (Ca)	[7440-70-2]	0,25	%w/w
ACCU-ASTM-P-0119-09	Chlorine (Cl)	[7782-50-5]	0,15	%w/w
ACCU-ASTM-P-0119-09	Copper (Cu)	[7440-50-8]	0,01	%w/w
ACCU-ASTM-P-0119-09	Magnesium (Mg)	[7439-95-4]	0,16	%w/w
ACCU-ASTM-P-0119-09	Phosphorus (P)	[7723-14-0]	0	%w/w
ACCU-ASTM-P-0119-09	Sulfur (S)	[7704-34-9]	0,35	%w/w
ACCU-ASTM-P-0119-09	Zinc (Zn)	[7440-66-6]	0,08	%w/w
ACCU-ASTM-P-0119-10	Calcium (Ca)	[7440-70-2]	0,11	%w/w
ACCU-ASTM-P-0119-10	Chlorine (Cl)	[7782-50-5]	0,15	%w/w
ACCU-ASTM-P-0119-10	Copper (Cu)	[7440-50-8]	0,04	%w/w
ACCU-ASTM-P-0119-10	Magnesium (Mg)	[7439-95-4]	0,005	%w/w
ACCU-ASTM-P-0119-10	Phosphorus (P)	[7723-14-0]	0,03	%w/w
ACCU-ASTM-P-0119-10	Sulfur (S)	[7704-34-9]	0,75	%w/w

ACCU-ASTM-P-0119-10	Zinc (Zn)	[7440-66-6]	0,15	%w/w
ACCU-ASTM-P-0119-11	Calcium (Ca)	[7440-70-2]	0,26	%w/w
ACCU-ASTM-P-0119-11	Chlorine (Cl)	[7782-50-5]	0,05	%w/w
ACCU-ASTM-P-0119-11	Copper (Cu)	[7440-50-8]	0	%w/w
ACCU-ASTM-P-0119-11	Magnesium (Mg)	[7439-95-4]	0	%w/w
ACCU-ASTM-P-0119-11	Phosphorus (P)	[7723-14-0]	0	%w/w
ACCU-ASTM-P-0119-11	Sulfur (S)	[7704-34-9]	0,75	%w/w
ACCU-ASTM-P-0119-11	Zinc (Zn)	[7440-66-6]	0	%w/w
ACCU-ASTM-P-0119-12	Calcium (Ca)	[7440-70-2]	0,2	%w/w
ACCU-ASTM-P-0119-12	Chlorine (Cl)	[7782-50-5]	0	%w/w
ACCU-ASTM-P-0119-12	Copper (Cu)	[7440-50-8]	0,005	%w/w
ACCU-ASTM-P-0119-12	Magnesium (Mg)	[7439-95-4]	0,14	%w/w
ACCU-ASTM-P-0119-12	Phosphorus (P)	[7723-14-0]	0,08	%w/w
ACCU-ASTM-P-0119-12	Sulfur (S)	[7704-34-9]	0,5	%w/w
ACCU-ASTM-P-0119-12	Zinc (Zn)	[7440-66-6]	0,08	%w/w
ACCU-ASTM-P-0119-13	Calcium (Ca)	[7440-70-2]	0	%w/w
ACCU-ASTM-P-0119-13	Chlorine (Cl)	[7782-50-5]	0	%w/w
ACCU-ASTM-P-0119-13	Copper (Cu)	[7440-50-8]	0,005	%w/w
ACCU-ASTM-P-0119-13	Magnesium (Mg)	[7439-95-4]	0,02	%w/w
ACCU-ASTM-P-0119-13	Phosphorus (P)	[7723-14-0]	0,02	%w/w
ACCU-ASTM-P-0119-13	Sulfur (S)	[7704-34-9]	0,2	%w/w
ACCU-ASTM-P-0119-13	Zinc (Zn)	[7440-66-6]	0,02	%w/w
ACCU-ASTM-P-0119-14	Calcium (Ca)	[7440-70-2]	0,07	%w/w
ACCU-ASTM-P-0119-14	Chlorine (Cl)	[7782-50-5]	0,15	%w/w
ACCU-ASTM-P-0119-14	Copper (Cu)	[7440-50-8]	0,02	%w/w
ACCU-ASTM-P-0119-14	Magnesium (Mg)	[7439-95-4]	0,08	%w/w
ACCU-ASTM-P-0119-14	Phosphorus (P)	[7723-14-0]	0,14	%w/w
ACCU-ASTM-P-0119-14	Sulfur (S)	[7704-34-9]	0,65	%w/w
ACCU-ASTM-P-0119-14	Zinc (Zn)	[7440-66-6]	0,15	%w/w
ACCU-ASTM-P-0119-15	Calcium (Ca)	[7440-70-2]	0,05	%w/w
ACCU-ASTM-P-0119-15	Chlorine (Cl)	[7782-50-5]	0	%w/w
ACCU-ASTM-P-0119-15	Copper (Cu)	[7440-50-8]	0	%w/w
ACCU-ASTM-P-0119-15	Magnesium (Mg)	[7439-95-4]	0	%w/w
ACCU-ASTM-P-0119-15	Phosphorus (P)	[7723-14-0]	0,15	%w/w
ACCU-ASTM-P-0119-15	Sulfur (S)	[7704-34-9]	0	%w/w
ACCU-ASTM-P-0119-15	Zinc (Zn)	[7440-66-6]	0	%w/w
ACCU-ASTM-P-0119-16	Calcium (Ca)	[7440-70-2]	0,4	%w/w
ACCU-ASTM-P-0119-16	Chlorine (Cl)	[7782-50-5]	0	%w/w

ACCU-ASTM-P-0119-16	Copper (Cu)	[7440-50-8]	0,001	%w/w
ACCU-ASTM-P-0119-16	Magnesium (Mg)	[7439-95-4]	0,08	%w/w
ACCU-ASTM-P-0119-16	Phosphorus (P)	[7723-14-0]	0	%w/w
ACCU-ASTM-P-0119-16	Sulfur (S)	[7704-34-9]	0,5	%w/w
ACCU-ASTM-P-0119-16	Zinc (Zn)	[7440-66-6]	0,02	%w/w
ACCU-ASTM-P-0119-17	Calcium (Ca)	[7440-70-2]	0,18	%w/w
ACCU-ASTM-P-0119-17	Chlorine (Cl)	[7782-50-5]	0,02	%w/w
ACCU-ASTM-P-0119-17	Copper (Cu)	[7440-50-8]	0,02	%w/w
ACCU-ASTM-P-0119-17	Magnesium (Mg)	[7439-95-4]	0	%w/w
ACCU-ASTM-P-0119-17	Phosphorus (P)	[7723-14-0]	0,02	%w/w
ACCU-ASTM-P-0119-17	Sulfur (S)	[7704-34-9]	0,6	%w/w
ACCU-ASTM-P-0119-17	Zinc (Zn)	[7440-66-6]	0,06	%w/w
ACCU-ASTM-P-0119-18	Calcium (Ca)	[7440-70-2]	0,4	%w/w
ACCU-ASTM-P-0119-18	Chlorine (Cl)	[7782-50-5]	0,01	%w/w
ACCU-ASTM-P-0119-18	Copper (Cu)	[7440-50-8]	0,001	%w/w
ACCU-ASTM-P-0119-18	Magnesium (Mg)	[7439-95-4]	0,01	%w/w
ACCU-ASTM-P-0119-18	Phosphorus (P)	[7723-14-0]	0,02	%w/w
ACCU-ASTM-P-0119-18	Sulfur (S)	[7704-34-9]	0	%w/w
ACCU-ASTM-P-0119-18	Zinc (Zn)	[7440-66-6]	0	%w/w
ACCU-ASTM-P-0119-19	Calcium (Ca)	[7440-70-2]	0,01	%w/w
ACCU-ASTM-P-0119-19	Chlorine (Cl)	[7782-50-5]	0,02	%w/w
ACCU-ASTM-P-0119-19	Copper (Cu)	[7440-50-8]	0,04	%w/w
ACCU-ASTM-P-0119-19	Magnesium (Mg)	[7439-95-4]	0,01	%w/w
ACCU-ASTM-P-0119-19	Phosphorus (P)	[7723-14-0]	0,02	%w/w
ACCU-ASTM-P-0119-19	Sulfur (S)	[7704-34-9]	0,2	%w/w
ACCU-ASTM-P-0119-19	Zinc (Zn)	[7440-66-6]	0,1	%w/w
ACCU-ASTM-P-0119-20	Calcium (Ca)	[7440-70-2]	0,05	%w/w
ACCU-ASTM-P-0119-20	Chlorine (Cl)	[7782-50-5]	0,005	%w/w
ACCU-ASTM-P-0119-20	Copper (Cu)	[7440-50-8]	0,05	%w/w
ACCU-ASTM-P-0119-20	Magnesium (Mg)	[7439-95-4]	0	%w/w
ACCU-ASTM-P-0119-20	Phosphorus (P)	[7723-14-0]	0,008	%w/w
ACCU-ASTM-P-0119-20	Sulfur (S)	[7704-34-9]	0	%w/w
ACCU-ASTM-P-0119-20	Zinc (Zn)	[7440-66-6]	0,12	%w/w
ACCU-ASTM-P-0119-21	Calcium (Ca)	[7440-70-2]	0,2	%w/w
ACCU-ASTM-P-0119-21	Chlorine (Cl)	[7782-50-5]	0,05	%w/w
ACCU-ASTM-P-0119-21	Copper (Cu)	[7440-50-8]	0,02	%w/w
ACCU-ASTM-P-0119-21	Magnesium (Mg)	[7439-95-4]	0,08	%w/w
ACCU-ASTM-P-0119-21	Phosphorus (P)	[7723-14-0]	0,05	%w/w

ACCU-ASTM-P-0119-21	Sulfur (S)	[7704-34-9]	0,275	%w/w
ACCU-ASTM-P-0119-21	Zinc (Zn)	[7440-66-6]	0,05	%w/w
ACCU-ASTM-P-0119-22	Calcium (Ca)	[7440-70-2]	0	%w/w
ACCU-ASTM-P-0119-22	Chlorine (Cl)	[7782-50-5]	0	%w/w
ACCU-ASTM-P-0119-22	Copper (Cu)	[7440-50-8]	0	%w/w
ACCU-ASTM-P-0119-22	Magnesium (Mg)	[7439-95-4]	0	%w/w
ACCU-ASTM-P-0119-22	Phosphorus (P)	[7723-14-0]	0	%w/w
ACCU-ASTM-P-0119-22	Sulfur (S)	[7704-34-9]	0	%w/w
ACCU-ASTM-P-0119-22	Zinc (Zn)	[7440-66-6]	0	%w/w