

QC Standard #2, 100 µg/mL in 2-5% Nitric Acid, tr Hydrofluoric acid

Art. ID ACCU-QCS-02-R1-5
Unit 500 mL
Deliverydetails ADR Limited Quantity (LQ) / AIR Dangerous Good (DG) UN3264 III E1

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Aluminium (Al)	[7429-90-5]	100	µg/mL		
	Barium (Ba)	[7440-39-3]	100	µg/mL		
	Boron (B)	[7440-42-8]	100	µg/mL		
	Potassium (K)	[7440-09-7]	100	µg/mL		
	Silicon (Si)	[7440-21-3]	100	µg/mL		
	Silver (Ag)	[7440-22-4]	100	µg/mL		
	Sodium (Na)	[7440-23-5]	100	µg/mL		