

XRF Drift Monitors Cement - 19 elements - 32mm

Art. ID XRFs-1201050-32MM
Unit each
Deliverydetails No Dangerous Good /not restricted

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Fluorine (F)	[7782-41-4]	4,9	%		
	Na ₂ O		2,7			
	MgO	[1309-48-4]	10,7			
	Al ₂ O ₃		8,7			
	SiO ₂		19,3			
	P ₂ O ₅		0,9			
	SO ₃		0,6			
	Chlorine (Cl)	[7782-50-5]	0,4			
	K ₂ O		1,4			
	CaO		26,9			
	TiO ₂		0,7			
	Cr ₂ O ₃		0,2			
	MnO		0,9			
	Fe ₂ O ₃		2			
	NiO		0,1			
	ZnO		0,2			
	As ₂ O ₃		0,2			
	SeO ₂		0,1			
	Bromine (Br)	[7726-95-6]	0,2			
	SrO		0,2			
	BaO		0,1			
	Thallium (Tl)	[7440-28-0]	0,1			
	PbO		0,1			