

ICP Multi-Element Standard - GENESIS Calibration Standard - 22 components in 2 %HNO₃ / 2% HCl/ tr. HF

Art. ID LM24-OP-4000138
Unit 100 mL
Deliverydetails ADR Limited Quantity (LQ) / AIR Dangerous Good (DG) UN3264 III E1

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Sulfur (S)	[7704-34-9]	50	mg/L		
	Cerium (Ce)	[7440-45-1]	10	mg/L		
	Copper (Cu)	[7440-50-8]	10	mg/L		
	Europium (Eu)	[7440-53-1]	10	mg/L		
	Iron (Fe)	[7439-89-6]	10	mg/L		
	Indium (In)	[7440-74-6]	10	mg/L		
	Potassium (K)	[7440-09-7]	10	mg/L		
	Nickel (Ni)	[7440-02-0]	10	mg/L		
	Phosphorus (P)	[7723-14-0]	10	mg/L		
	Silicon (Si)	[7440-21-3]	10	mg/L		
	Titanium (Ti)	[7440-32-6]	10	mg/L		
	Vanadium (V)	[7440-62-2]	10	mg/L		
	Yttrium (Y)	[7440-65-5]	10	mg/L		
	Zirconium (Zr)	[7440-67-7]	10	mg/L		
	Manganese (Mn)	[7439-96-5]	5	mg/L		
	Molybdenum (Mo)	[7439-98-7]	5	mg/L		
	Sodium (Na)	[7440-23-5]	5	mg/L		
	Scandium (Sc)	[7440-20-2]	5	mg/L		
	Beryllium (Be)	[7440-41-7]	2	mg/L		
	Lithium (Li)	[7439-93-2]	2	mg/L		
	Strontium (Sr)	[7440-24-6]	2	mg/L		
	Calcium (Ca)	[7440-70-2]	1	mg/L		