

Column Test Mixture - 17 components (ASTM D2887-06a)

Art. ID LM24-OP-4001914
Unit 1 mL
Deliverydetails De Minimis UN1993 II E2

Description

Validity: 24 months

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	n-Hexane (C6)	[110-54-3]	6	%		
	n-Heptane (C7)	[142-82-5]	6	%		
	n-Octane (C8)	[111-65-9]	8	%		
	n-Nonane (C9)	[111-84-2]	8	%		
	n-Decane (C10)	[124-18-5]	12	%		
	n-Undecane (C11)	[1120-21-4]	12	%		
	n-Dodecane (C12)	[112-40-3]	12	%		
	n-Tetradecane (C14)	[629-59-4]	12	%		
	n-Hexadecane (C16)	[544-76-3]	10	%		
	n-Octadecane (C18)	[593-45-3]	5	%		
	n-Eicosane (C20)	[112-95-8]	2	%		
	n-Tetracosane (C24)	[646-31-1]	2	%		
	n-Octacosane (C28)	[630-02-4]	1	%		
	n-Dotriacontane (C32)	[544-85-4]	1	%		
	n-Hexatriacontane (C36)	[630-06-8]	1	%		
	n-Tetracontane (C40)	[4181-95-7]	1	%		
	n-Tetratetracontane (C44)	[7098-22-8]	1	%		