

Nickel base, RM, 200, RoHS Monitor, center has small crack, not for XRF, Disc 31 mm x 2 mm

Art. ID IARM-189A-X
Unit disc
Deliverydetails No Dangerous Good /not restricted

| Text/Information | Analyte/Parameter | CAS number | Concentration/Value | Unit | Method | Source |
|------------------|-------------------|--------------|---------------------|------|--------|--------|
| | Silver (Ag) | [7440-22-4] | 0,00024 | % | | |
| | Aluminium (Al) | [7429-90-5] | 0,0044 | % | | |
| | Arsenic (As) | [7440-38-2] | 0,00007 | % | | |
| | Boron (B) | [7440-42-8] | (<0,0005) | % | | |
| | Beryllium (Be) | [7440-41-7] | (<0,0001) | % | | |
| | Bismuth (Bi) | [7440-69-9] | 0,00026 | % | | |
| | Carbon (C) | [7440-44-0] | 0,0023 | % | | |
| | Calcium (Ca) | [7440-70-2] | <0,0010 | % | | |
| | Cadmium (Cd) | [7440-43-9] | 0,00008 | % | | |
| | Cobalt (Co) | [7440-48-4] | 0,00031 | % | | |
| | Chromium (Cr) | [7440-47-3] | ~0,001 | % | | |
| | Copper (Cu) | [7440-50-8] | 0,0009 | % | | |
| | Iron (Fe) | [7439-89-6] | 0,0038 | % | | |
| | Gallium (Ga) | [7440-55-3] | <0,00005 | % | | |
| | Magnesium (Mg) | [7439-95-4] | ~0,0008 | % | | |
| | Manganese (Mn) | [7439-96-5] | 0,00019 | % | | |
| | Molybdenum (Mo) | [7439-98-7] | (<0,0001) | % | | |
| | Nitrogen (N) | [7727-37-9] | ~0,0001 | % | | |
| | Nickel (Ni) | [7440-02-0] | rem | % | | |
| | Oxygen (O) | [7782-44-7] | ~0,0018 | % | | |
| | Phosphorus (P) | [7723-14-0] | 0,00037 | % | | |
| | Lead (Pb) | [7439-92-1] | 0,00029 | % | | |
| | Sulfur (S) | [7704-34-9] | 0,00018 | % | | |
| | Antimony (Sb) | [7440-36-0] | 0,00039 | % | | |
| | Selenium (Se) | [7782-49-2] | 0,00021 | % | | |
| | Silicon (Si) | [7440-21-3] | 0,0019 | % | | |
| | Tin (Sn) | [7440-31-5] | 0,00022 | % | | |
| | Tellurium (Te) | [13494-80-9] | 0,00017 | % | | |
| | Titanium (Ti) | [7440-32-6] | ~0,0003 | % | | |
| | Thallium (Tl) | [7440-28-0] | 0,00023 | % | | |

| | | | |
|--------------|-------------|------------|---|
| Vanadium (V) | [7440-62-2] | (<0,00005) | % |
| Zinc (Zn) | [7440-66-6] | 0,00028 | % |