

Out of Stock - Product is not available at this time - Portland cement - Turbidity and Fineness

Art. ID NIST-46h
Unit 10 x 5 g
Deliverydetails No Dangerous Good /not restricted

Description

This Standard Reference Material (SRM®) is intended for use in calibrating fineness testing equipment according to ASTM Standard Methods. A unit of NIST-46h contains 10 glass vials of approximately 5 g of powdered cement. Each vial has a plastic snap off cap and is contained in a sealed foil bag. *Standard Test Method for Fineness of Hydraulic Cement by the 45 µm (No. 325) Sieve. Certified values /// Sample value(s) - please ask for current certificate.

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Sieve Residue (45 µm residue)		7.43 ± 0.79	%	ASTM C 430 -96(2003)*	