

Multielement aqueous calibration solution in 10% HNO₃

Art. ID AP-CZ9086MN1
Unit 100 mL/HDPE
Deliverydetails ADR Limited Quantity (LQ) / AIR Dangerous Good (DG) UN2031 II E2

| Text/Information | Analyte/Parameter | CAS number | Concentration/Value | Unit | Method | Source |
|------------------|-------------------|-------------|---------------------|------|--------|--------|
| | Aluminium (Al) | [7429-90-5] | 100 ± 0.2 | mg/L | | |
| | Thallium (Tl) | [7440-28-0] | 100 ± 0.2 | mg/L | | |
| | Boron (B) | [7440-42-8] | 20 ± 0.04 | mg/L | | |
| | Cadmium (Cd) | [7440-43-9] | 20 ± 0.04 | mg/L | | |
| | Chromium (Cr) | [7440-47-3] | 20 ± 0.04 | mg/L | | |
| | Copper (Cu) | [7440-50-8] | 20 ± 0.04 | mg/L | | |
| | Iron (Fe) | [7439-89-6] | 20 ± 0.04 | mg/L | | |
| | Barium (Ba) | [7440-39-3] | 5 ± 0.01 | mg/L | | |
| | Selenium (Se) | [7782-49-2] | 5 ± 0.01 | mg/L | | |
| | Beryllium (Be) | [7440-41-7] | 2 ± 0.004 | mg/L | | |
| | Cobalt (Co) | [7440-48-4] | 50 ± 0.1 | mg/L | | |
| | Nickel (Ni) | [7440-02-0] | 50 ± 0.1 | mg/L | | |
| | Vanadium (V) | [7440-62-2] | 50 ± 0.1 | mg/L | | |
| | Zinc (Zn) | [7440-66-6] | 50 ± 0.1 | mg/L | | |
| | Manganese (Mn) | [7439-96-5] | 10 ± 0.02 | mg/L | | |
| | Lead (Pb) | [7439-92-1] | 200 ± 0.4 | mg/L | | |