

### **Multielement (ion) calibration solution in Water**

Art. ID AP-CZ9102MH1IC  
Unit 100 mL/HDPE  
Deliverydetails No Dangerous Good /not restricted

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	Bromide (Br-)		100 ± 0.2	mg/L		
	Chloride (Cl-)		100 ± 0.2	mg/L		
	Fluoride (F-)		100 ± 0.2	mg/L		
	Nitrate (NO <sub>3</sub> -)		100 ± 0.2	mg/L		
	Phosphate (PO <sub>4</sub> <sup>3-</sup> )		100 ± 0.2	mg/L		
	Sulfate (SO <sub>4</sub> <sup>2-</sup> )		100 ± 0.2	mg/L		