

1408.8 μ S/cm Conductivity - 1 L

Art. ID	ERA-01620
Unit	1 L
Deliverydetails	No Dangerous Good

Description

Conductivity calibration or verification standard prepared from KCl and ultrapure water at 1408.8 uS/cm - 1L 6 month expiration