

### **1408.8 $\mu$ S/cm Conductivity - 1 L**

Art. ID	ERA-01620
Unit	1 L
Deliverydetails	No Dangerous Good

#### Description

Conductivity calibration or verification standard prepared from KCl and ultrapure water at 1408.8 uS/cm - 1L 6 month expiration