

**30100  $\mu$ S/cm Conductivity-250 mL**

Art. ID	ERA-01645
Unit	250 mL
Deliverydetails	No Dangerous Good

**Description**

Conductivity calibration or verification standard prepared from KCl and ultrapure water at 30100 uS/cm - 250mL 6 month expiration