

**100,000  $\mu$ S/cm Conductivity - 500 mL**

Art. ID	ERA-01655
Unit	500 mL
Deliverydetails	No Dangerous Good

**Description**

Conductivity calibration / verification CRM prepared from KCl and ultrapure water at 100,000  $\mu$ S/cm - 500mL. 6 Month expiration.