

500,000 μ S/cm Conductivity - 500 mL

Art. ID	ERA-01668
Unit	500 mL
Deliverydetails	ADR Limited Quantity (LQ) / AIR Dangerous Good (DG) UN1789 III E1

Description

Conductivity calibration / verification CRM prepared from KCl and ultrapure water at 500,000 μ S/cm - 500 mL 6 month expiration