

25 μ S/cm Conductivity - Sievers 500 - 30 mL

Art. ID	ERA-02750
Unit	30 mL
Deliverydetails	No Dangerous Good

Description

Conductivity calibration or verification standard prepared from HCl and ultrapure water at 25 uS/cm. Packaged in 30 mL vials specifically for use in Sievers Model 500 instruments.