

TT - OI Analytical TOC Validation Set - Aurora

Art. ID	ERA-19007
Unit	Set
Deliverydetails	No Dangerous Good

Description

Set includes nineteen 40 mL vials, (6) Water blank, (1) 0.500 ppm NIST KHP, (3) 1.00 ppm NIST KHP, (1) 5.00 ppm NIST KHP, (1) 10.00 ppm NIST KHP, (1) 25.0ppm NIST KHP, (1) 5 ppm NIST KHP/50 ppm NaHCO₃, (4) 0.500 mg C/L from USP Sucrose RS, (1) 0.500 mg C/L from USP 1,4-benzoquinone RSS.