

PAT700 JP Calibration Stds Kit #2

Art. ID	ERA-19203J
Unit	Kit
Deliverydetails	No Dangerous Good

Description

Set of 4 bottles includes (1) Blank, (1) 0.500 mg/L C NIST KHP, (1) 0.750 mg/L C NIST KHP, (1) 1.250 mg/L C NIST KHP in 60 ml bottles with RFID Tags for JP Compliance and most current instrument firmware. Use for PAT700.