

Anatel A-643 Validation Protocol Kit -60 mL vials

Art. ID	ERA-19230
Unit	60 mL
Deliverydetails	No Dangerous Good

Description

Set includes 38 vials, (10) Blanks, (3) 0.250 mg/L C NIST Sucrose, (5) 0.500 mg/LC NIST Sucrose, (3) 0.750 mg/L C NIST Sucrose, (1) 100 µS/cm Conductivity Std., (4) USP System Suitability Sets. Replaces reference materials portion of FG5017701 - lamp.