

## **PAT700 Validation Protocol Reagent Kit**

Art. ID	ERA-19232
Unit	Kit
Deliverydetails	No Dangerous Good

### Description

(5) Conductivity Solution Kits (2) Calibration Standard Kits (1) Validation Control Kit (1) System Suitability Set (1) Excursion with Validation Kit (14) TOC Calibration Blanks (1) 0.250 mg C/L from NIST Sucrose (1) 0.750 mg C/L from NIST Sucrose (1) Reagent Water Rw (1) 0.500 mg C/L from USP 1,4-benzoquinone Rss (1) Excursion Bottle  
/// Waters ERA will continue to guarantee a minimum shelf life of 30 days from date of shipment /// Product support the compliance for both USP <643> and EP 2.2.44