

## **Sievers Model 800 Auto Sampler Calibration Set**

Art. ID	ERA-19500
Unit	Set
Deliverydetails	No Dangerous Good

### Description

Set includes twelve 40 mL vials, (4) blanks, (4) 25.0 mg/L C NIST KHP, (4) 25.0 mg/L IC NIST NaHCO<sub>3</sub>. Replaces Sievers® Part STD 31002.