

Sievers® Model 900 Autoreagents Calibration Set

Art. ID	ERA-19611
Unit	Set
Deliverydetails	No Dangerous Good

Description

Set includes four 40 mL vials, (1) blank, (1) 25.0 mg/L C NIST KHP, (1) 25 mg/L IC NIST NaHCO₃, (1) 50.0 mg/L C from Nicotinamide. Replaces Sievers® Part STD 90036.