

Sievers 900 Autoreagents Verification Set

Art. ID	ERA-19614
Unit	Set
Deliverydetails	No Dangerous Good

Description

Sievers Model 900 Verification Set (includes (1) calibration blank, (1) 25.0 ppm NIST Sucrose, (1) 50.0 ppm NIST Sucrose, (1) 50.0 ppm IC as NaHCO₃ - all 40 mL