

Sievers Model 900 Autoreagents Calibration & Verification Set

Art. ID	ERA-19616
Unit	Set
Deliverydetails	No Dangerous Good

Description

Set Includes eight 40 mL vials, (1) Blank, (1) 25.0 mg/L C NIST KHP, (1) 25.0 mg/L IC NaHCO₃, (1) 50 mg/L C from Nicotinamide, (1) Blank, (1) 25.0 mg/L C NIST Sucrose, (1) 50.0 mg/L C NIST Sucrose, (1) 50.0 Mg/L IC NAHCO₃ - All 40 mL. Replaces Sievers® CSTD90028.