

Sievers® Model 900 Validation Set with Calibration and Verification

Art. ID	ERA-19617
Unit	Set
Deliverydetails	No Dangerous Good

Description

Set Includes (28) 40 mL vials, (2) Blank, (2) 1.0 mg/L C NIST KHP, (2) 1.0 mg/L IC NaHCO₃, (1) Blank, (1) 0.5 mg/L C NIST Sucrose, (1) 0.5 mg/L IC NA₂CO₃, (2) Reagent Water Rw, (2) 0.500 mg/L C from USP Sucrose Rs and (2) 0.50 mg/L C from USP 1,4-benzoquinone Rss, (1) Blank, (1) 0.25 mg/L C NIST Sucrose, (1) 0.50 mg/L C NIST Sucrose, (1) 0.75 mg/L C NIST Sucrose, (1) Reagent Waters Rw, (1) 0.50 mg/L C from USP Sucrose Rs, (1) Blank, (1) .50mg/L C from Nicotinamide, (1) .50 mg/L C from NIST KHP, (1) Reagent Water Rw, (1) .50 mg/L C from USP Sucrose Rs, and (1) .50 mg/L C from USP 1,4-Benzoquinone Rss. Replaces Sievers® CSTD90025.