

Validex™ Water Count Reference Standard NIST Traceable Mean Diameter

Art. ID	FIS-11858273
Unit	500 mL
Deliverydetails	No Dangerous Good /not restricted

Description

Validex™ Water Count Reference Standard consist of microspheres suspended in water at a concentration fo 1000 particles per milliliter (mL). These single-use bottles are designed to verify the precision of particle counters. The certified mean diameter is certified as traceable to the Standard Meter throught the National Institute of Standards and Techenology (NIST) /// Please note that this item is not returnable

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
Certified value, nomina I	Mean Diameter		10	µm		