

500,000 μ S/cm Conductivity - 125 mL

Art. ID	ERA-01667
Unit	125 mL
Deliverydetails	ADR Limited Quantity (LQ) / AIR Dangerous Good (DG) UN1789 III E1

Description

Conductivity calibration / verification CRM prepared from KCl and ultrapure water at 500,000 μ S/cm - 125 mL - 6 month expiration