

Conductivity Validation Kit - 125 mL

Art. ID	ERA-02850
Unit	125 mL

Description

125 mL Kit includes four bottles, (1) 25 $\mu\text{S}/\text{cm}$ in Water, (1) 100 $\mu\text{S}/\text{cm}$ in Water, (1) 146.93 $\mu\text{S}/\text{cm}$ in Water (Solution D) and (1) Reagent Blank for μSe with Solution D.