

LL System Suitability Set - 40 mL

Art. ID ERA-18040
Unit Set

Description

Low Level Water System Suitability - Set of three vials includes (1) Reagent Water Rw, (1) 0.3 mg/L C from μ SP Sucrose Rs, (1) 0.3 mg/L C from μ SP 1,4-Benzoquinone Rss in 40 mL glass vials. μ Se for the following instruments: Sievers 800, 900, Analytik Jena, Shimadzu UV/Persulfate, & Teledyne Tekmar.