

LL System Suitability Rss - 40 mL

Art. ID ERA-18065
Unit 40 mL

Description

Low Level Water System Suitability Rss - Component of 18061 Set includes (1) 0.3 mg/L C from μ SP
1,4-Benzoquinone Rss in 40 mL glass vial. μ Se for the following instruments: Sievers 800, 900, Analytik Jena,
Shimadzu UV/Persulfate, & Teledyne Tekmar.