

## **LL System Suitability Rss - 40 mL**

Art. ID                      ERA-18065  
Unit                         40 mL

### Description

Low Level Water System Suitability Rss - Component of 18061 Set includes (1) 0.3 mg/L C from  $\mu$ SP 1,4-Benzoquinone Rss in 40 mL glass vial.  $\mu$ Se for the following instruments: Sievers 800, 900, Analytik Jena, Shimadzu UV/Persulfate, & Teledyne Tekmar.