

ASTM Method D2887 Calibration Solution: 17 components in Chloroform

Art. ID LM24-OP-4011632
Unit 1 mL
Deliverydetails De Minimis UN1888 III E1

Description

Validity: 12 months

| Text/Information | Analyte/Parameter | CAS number | Concentration/Value | Unit | Method | Source |
|------------------|---------------------|-------------|---------------------|------|--------|--------|
| | n-Decane | [124-18-5] | 1200 | mg/L | | |
| | n-Dodecane | [112-40-3] | 1200 | mg/L | | |
| | n-Dotriacontane | [544-85-4] | 100 | mg/L | | |
| | n-Eicosane | [112-95-8] | 200 | mg/L | | |
| | n-Heptane | [142-82-5] | 600 | mg/L | | |
| | n-Hexadecane | [544-76-3] | 1000 | mg/L | | |
| | n-Hexane | [110-54-3] | 600 | mg/L | | |
| | n-Hexatriacontane | [630-06-8] | 100 | mg/L | | |
| | n-Nonane | [111-84-2] | 800 | mg/L | | |
| | n-Octadecane | [593-45-3] | 500 | mg/L | | |
| | n-Octane | [111-65-9] | 800 | mg/L | | |
| | n-Tetradecane | [629-59-4] | 1200 | mg/L | | |
| | n-Tetratetracontane | [7098-22-8] | 100 | mg/L | | |
| | n-Undecane | [1120-21-4] | 1200 | mg/L | | |
| | n-Octacosane | [630-02-4] | 100 | mg/L | | |
| | n-Tetracontane | [4181-95-7] | 100 | mg/L | | |
| | n-Tetracosane | [646-31-1] | 200 | mg/L | | |