

ASTM Method D2887 Calibration Solution: 17 components in Chloroform

Art. ID LM24-OP-4011632
Unit 1 mL
Deliverydetails De Minimis UN1888 III E1

Description

Validity: 12 months

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
	n-Decane	[124-18-5]	1200	mg/L		
	n-Dodecane	[112-40-3]	1200	mg/L		
	n-Dotriacontane	[544-85-4]	100	mg/L		
	n-Eicosane	[112-95-8]	200	mg/L		
	n-Heptane	[142-82-5]	600	mg/L		
	n-Hexadecane	[544-76-3]	1000	mg/L		
	n-Hexane	[110-54-3]	600	mg/L		
	n-Hexatriacontane	[630-06-8]	100	mg/L		
	n-Nonane	[111-84-2]	800	mg/L		
	n-Octadecane	[593-45-3]	500	mg/L		
	n-Octane	[111-65-9]	800	mg/L		
	n-Tetradecane	[629-59-4]	1200	mg/L		
	n-Tetratetracontane	[7098-22-8]	100	mg/L		
	n-Undecane	[1120-21-4]	1200	mg/L		
	n-Octacosane	[630-02-4]	100	mg/L		
	n-Tetracontane	[4181-95-7]	100	mg/L		
	n-Tetracosane	[646-31-1]	200	mg/L		