

Method 200,7 Calibration Std Set - Revised

Art. ID ACCU-M-200.7-R-1-SET
Unit 5 x 100 mL
Deliverydetails ADR Limited Quantity (LQ) / AIR Dangerous Good (DG)

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
ACCU-M-200.7-01-1	Antimony (Sb)	[7440-36-0]	50	µg/mL		
ACCU-M-200.7-01-1	Arsenic (As)	[7440-38-2]	100	µg/mL		
ACCU-M-200.7-01-1	Barium (Ba)	[7440-39-3]	10	µg/mL		
ACCU-M-200.7-01-1	Boron (B)	[7440-42-8]	20	µg/mL		
ACCU-M-200.7-01-1	Cadmium (Cd)	[7440-43-9]	20	µg/mL		
ACCU-M-200.7-01-1	Calcium (Ca)	[7440-70-2]	100	µg/mL		
ACCU-M-200.7-01-1	Copper (Cu)	[7440-50-8]	20	µg/mL		
ACCU-M-200.7-01-1	Manganese (Mn)	[7439-96-5]	20	µg/mL		
ACCU-M-200.7-01-1	Selenium (Se)	[7782-49-2]	50	µg/mL		
ACCU-M-200.7-01-1	Silver (Ag)	[7440-22-4]	5	µg/mL		
ACCU-M-200.7-02R-1	Lithium (Li)	[7439-93-2]	50	µg/mL		
ACCU-M-200.7-02R-1	Molybdenum (Mo)	[7439-98-7]	100	µg/mL		
ACCU-M-200.7-02R-1	Potassium (K)	[7440-09-7]	200	µg/mL		
ACCU-M-200.7-02R-1	Sodium (Na)	[7440-23-5]	100	µg/mL		
ACCU-M-200.7-02R-1	Strontium (Sr)	[7440-24-6]	10	µg/mL		
ACCU-M-200.7-02R-1	Titanium (Ti)	[7440-32-6]	100	µg/mL		
ACCU-M-200.7-03R-1	Cerium (Ce)	[7440-45-1]	20	µg/mL		
ACCU-M-200.7-03R-1	Cobalt (Co)	[7440-48-4]	20	µg/mL		
ACCU-M-200.7-03R-1	Phosphorus (P)	[7723-14-0]	100	µg/mL		
ACCU-M-200.7-03R-1	Vanadium (V)	[7440-62-2]	20	µg/mL		
ACCU-M-200.7-04-1	Aluminium (Al)	[7429-90-5]	100	µg/mL		
ACCU-M-200.7-04-1	Chromium (Cr)	[7440-47-3]	50	µg/mL		
ACCU-M-200.7-04-1	Silicon (Si)	[7440-21-3]	100	µg/mL		
ACCU-M-200.7-04-1	Tin (Sn)	[7440-31-5]	40	µg/mL		
ACCU-M-200.7-04-1	Zinc (Zn)	[7440-66-6]	50	µg/mL		
ACCU-M-200.7-05-1	Beryllium (Be)	[7440-41-7]	10	µg/mL		
ACCU-M-200.7-05-1	Iron (Fe)	[7439-89-6]	100	µg/mL		
ACCU-M-200.7-05-1	Lead (Pb)	[7439-92-1]	100	µg/mL		
ACCU-M-200.7-05-1	Magnesium (Mg)	[7439-95-4]	100	µg/mL		
ACCU-M-200.7-05-1	Nickel (Ni)	[7440-02-0]	20	µg/mL		
ACCU-M-200.7-05-1	Thallium (Tl)	[7440-28-0]	50	µg/mL		

