

Setting-Up sample for spectrometric analysis of low allowed steels - Ø 50 mm x 42 mm

Art. ID BAM-SUS-1 R
Unit cylinder
Deliverydetails No Dangerous Good /not restricted

| Text/Information | Analyte/Parameter | CAS number | Concentration/Value | Unit | Method | Source |
|------------------|-------------------|-------------|---------------------|------|--------|--------|
| | Carbon (C) | [7440-44-0] | 0,9 | % | | |
| | Silicon (Si) | [7440-21-3] | 0,8 | % | | |
| | Manganese (Mn) | [7439-96-5] | 1,1 | % | | |
| | Phosphorus (P) | [7723-14-0] | 0,02 | % | | |
| | Sulfur (S) | [7704-34-9] | 0,017 | % | | |
| | Chromium (Cr) | [7440-47-3] | 1,7 | % | | |
| | Molybdenum (Mo) | [7439-98-7] | 0,9 | % | | |
| | Nickel (Ni) | [7440-02-0] | 2,9 | % | | |
| | Vanadium (V) | [7440-62-2] | 0,5 | % | | |
| | Tungsten (W) | [7440-33-7] | 0,7 | % | | |
| | Copper (Cu) | [7440-50-8] | 0,7 | % | | |
| | Cobalt (Co) | [7440-48-4] | 0,3 | % | | |
| | Niobium (Nb) | [7440-03-1] | 0,55 | % | | |