

**ROHS-PVC9-5ED-31 Plastic Polyvinyl Chloride PVC Calibration Standards for WEEE
RoHS Requirements, Disc Ø 31 mm x thickness on request, 9 standards per set (Includes
1x ASI-ROHS-PVC9-5ED-QC-31)**

Art. ID	ASI-ROHS-PVC9-5ED-31
Unit	Set of 10 discs
Deliverydetails	No Dangerous Good /not restricted

Description

Plastic Polyvinyl Chloride (PVC) Calibration Standards for WEEE RoHS Requirements, Disc, 9 standards per set, 5 elements per standard, 31 mm x thickness on request, Br, Cd, Cr, Hg, Pb. Includes 1 ASI-ROHS-PVC9-5ED-QC-31 Check Standard.

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
Sample 1	Bromine (Br)	[7726-95-6]	0	%		
Sample 1	Cadmium (Cd)	[7440-43-9]	0	%		
Sample 1	Chromium (Cr)	[7440-47-3]	0	%		
Sample 1	Mercury (Hg)	[7439-97-6]	0	%		
Sample 1	Lead (Pb)	[7439-92-1]	0	%		
Sample2	Bromine (Br)	[7726-95-6]	0,005	%		
Sample 2	Cadmium (Cd)	[7440-43-9]	0,0025	%		
Sample2	Chromium (Cr)	[7440-47-3]	0,005	%		
Sample2	Mercury (Hg)	[7439-97-6]	0,01	%		
Sample2	Lead (Pb)	[7439-92-1]	0,1	%		
Sample 3	Bromine (Br)	[7726-95-6]	0,04	%		
Sample 3	Cadmium (Cd)	[7440-43-9]	0,01	%		
Sample 3	Chromium (Cr)	[7440-47-3]	0,075	%		
Sample 3	Mercury (Hg)	[7439-97-6]	0,0075	%		
Sample 3	Lead (Pb)	[7439-92-1]	0,025	%		
Sample 4	Bromine (Br)	[7726-95-6]	0,01	%		
Sample 4	Cadmium (Cd)	[7440-43-9]	0,0125	%		
Sample 4	Chromium (Cr)	[7440-47-3]	0,125	%		
Sample 4	Mercury (Hg)	[7439-97-6]	0,05	%		
Sample 4	Lead (Pb)	[7439-92-1]	0,05	%		
Sample 5	Bromine (Br)	[7726-95-6]	0,025	%		
Sample 5	Cadmium (Cd)	[7440-43-9]	0,0075	%		
Sample 5	Chromium (Cr)	[7440-47-3]	0,1	%		
Sample 5	Mercury (Hg)	[7439-97-6]	0,025	%		
Sample 5	Lead (Pb)	[7439-92-1]	0,125	%		
Sample 6	Bromine (Br)	[7726-95-6]	0,05	%		
Sample 6	Cadmium (Cd)	[7440-43-9]	0,001	%		

Sample 6	Chromium (Cr)	[7440-47-3]	0,065	%
Sample 6	Mercury (Hg)	[7439-97-6]	0,08	%
Sample 6	Lead (Pb)	[7439-92-1]	0,075	%
Sample 7	Bromine (Br)	[7726-95-6]	0,02	%
Sample 7	Cadmium (Cd)	[7440-43-9]	0,0005	%
Sample 7	Chromium (Cr)	[7440-47-3]	0,025	%
Sample 7	Mercury (Hg)	[7439-97-6]	0,1	%
Sample 7	Lead (Pb)	[7439-92-1]	0,01	%
Sample 8	Bromine (Br)	[7726-95-6]	0,03	%
Sample 8	Cadmium (Cd)	[7440-43-9]	0,005	%
Sample 8	Chromium (Cr)	[7440-47-3]	0,05	%
Sample 8	Mercury (Hg)	[7439-97-6]	0,003	%
Sample 8	Lead (Pb)	[7439-92-1]	0,05	%
Sample 9	Bromine (Br)	[7726-95-6]	0,005	%
Sample 9	Cadmium (Cd)	[7440-43-9]	0,015	%
Sample 9	Chromium (Cr)	[7440-47-3]	0,01	%
Sample 9	Mercury (Hg)	[7439-97-6]	0,12	%
Sample 9	Lead (Pb)	[7439-92-1]	0,035	%
QC Sample 1: ASI-ROHS-F Bromine (Br)		[7726-95-6]	0,025	%
VC9-5ED-QC-31				
QC Sample 1: ASI-ROHS-F Cadmium (Cd)		[7440-43-9]	0,005	%
VC9-5ED-QC-31				
QC Sample 1: ASI-ROHS-F Chromium (Cr)		[7440-47-3]	0,05	%
VC9-5ED-QC-31				
QC Sample 1: ASI-ROHS-F Mercury (Hg)		[7439-97-6]	0,05	%
VC9-5ED-QC-31				
QC Sample 1: ASI-ROHS-F Lead (Pb)		[7439-92-1]	0,05	%
VC9-5ED-QC-31				