

## **Stainless Steel Grade 304L Disc 38 mm D x 19 mm**

Art. ID	BS-304C
Unit	disc
Deliverydetails	No Dangerous Good /not restricted

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
Certified Value	Aluminium (Al)	[7429-90-5]	0,0022 ± 0,0007	%		
Certified Value	Arsenic (As)	[7440-38-2]	0,0047 ± 0,0005	%		
Certified Value	Boron (B)	[7440-42-8]	0,0005 ± 0,0002	%		
Certified Value	Carbon (C)	[7440-44-0]	0,024 ± 0,003	%		
Certified Value	Cobalt (Co)	[7440-48-4]	0,229 ± 0,009	%		
Certified Value	Chromium (Cr)	[7440-47-3]	18,2 ± 0,2	%		
Certified Value	Copper (Cu)	[7440-50-8]	0,524 ± 0,009	%		
Certified Value	Iron (Fe)	[7439-89-6]	70,1 ± 0,2	%		
Certified Value	Hydrogen (H)	[1333-74-0]	0,0004 ± 0,0002	%		
Certified Value	Manganese (Mn)	[7439-96-5]	1,37 ± 0,02	%		
Certified Value	Molybdenum (Mo)	[7439-98-7]	0,419 ± 0,009	%		
Certified Value	Nitrogen (N)	[7727-37-9]	0,077 ± 0,004	%		
Certified Value	Nickel (Ni)	[7440-02-0]	8,53 ± 0,08	%		
Certified Value	Oxygen (O)	[7782-44-7]	0,005 ± 0,0005	%		
Certified Value	Phosphorus (P)	[7723-14-0]	0,034 ± 0,003	%		
Certified Value	Sulfur (S)	[7704-34-9]	0,027 ± 0,003	%		
Certified Value	Silicon (Si)	[7440-21-3]	0,302 ± 0,009	%		
Certified Value	Tin (Sn)	[7440-31-5]	0,0119 ± 0,0009	%		
Certified Value	Vanadium (V)	[7440-62-2]	0,083 ± 0,004	%		
Certified Value	Tungsten (W)	[7440-33-7]	0,043 ± 0,004	%		
Reference Value	Calcium (Ca)	[7440-70-2]	0,0002 ± 0,0001	%		
Reference Value	Magnesium (Mg)	[7439-95-4]	<0,005	%		
Reference Value	Niobium (Nb)	[7440-03-1]	0,009 ± 0,004	%		
Reference Value	Antimony (Sb)	[7440-36-0]	<0,05	%		
Reference Value	Titanium (Ti)	[7440-32-6]	0,0024 ± 0,0014	%		
Reference Value	Zinc (Zn)	[7440-66-6]	<0,005	%		
Reference Value	Zirconium (Zr)	[7440-67-7]	0,002 ± 0,001	%		