

Low-Energy Charpy V-Notch Specimens (Self-Verification, 21 °C, 8 mm Striker)

Art. ID	NIST-2562
Unit	set
Deliverydetails	No Dangerous Good /not restricted

Description

The certified values delivered by this Standard Reference Material (SRM) are intended for the verification of Charpy machines equipped with an 8 mm striker, in accordance with the current ASTM Standard E23 and the current ISO Standard 148-2. A unit of SRM 2562 consists of a set of Charpy specimens needed to perform one in-house verification (self-verification). Specimens are made from 4340 alloy steel. The bars are finished to length, stamped, heat treated, and machined in SRM specimen lots of approximately 1900 specimens. Each specimen has a lot number and an identification number (three or four digits).

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
Please ask for current certificate						