

Thromboplastin, rabbit

Art. ID	ERM-AD149
Unit	ampoule
Deliverydetails	No Dangerous Good /not restricted

Description

European Reference Material ERM® -AD149 was originally certified as BCR-149S. It was produced and certified under the responsibility of the IRMM according to the principles laid down in the technical guidelines of the European Reference Materials® co-operation agreement between BAM-IRMM-LGC. The reference material AD149 was measured against RBT/90. Coagulation times of normal and patient plasmas measured with both thromboplastins were compared. A linear relationship was established between the logarithms of the values obtained for the coagulation times/s in both systems.

Text/Information	Analyte/Parameter	CAS number	Concentration/Value	Unit	Method	Source
Parameters of the regression line	Prothrombin time slope		1,257 ± 0,013			
Parameters of the regression line	Prothrombin time intercept		(- 0,242 ± 0,019)			